

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091119\
 Data File : PQ043503.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2019 01:57
 Operator : SM\AJ
 Sample : K3591-08
 Misc : AR1254 LOQ 50 PPB
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOQ-WATER-02-QT3-2019

Manual Integrations
 APPROVED

Ankita
 9/12/2019 12:16:06 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 12 02:56:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 12 02:33:14 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.693	4.782	37068866	13869709	23.265m	22.898
2)	SA Decachlor...	12.080	10.552	139.9E6	69938316	22.175	22.329
Target Compounds							
26)	L6 AR-1254-1	8.095	7.189	6744209	3648719	93.941	80.662
27)	L6 AR-1254-2	8.333	7.356	11376351	3642677	92.528	82.654
28)	L6 AR-1254-3	8.726	7.804	13830367	6457229	95.095	86.566
29)	L6 AR-1254-4	9.028	8.053	10999747	4543993	85.462	82.109
30)	L6 AR-1254-5	9.466	8.499	12796885	5781361	84.265	70.939

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091119\
Data File : PQ043503.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2019 01:57
Operator : SM\AJ
Sample : K3591-08
Misc : AR1254 LOQ 50 PPB
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2019

Manual Integrations
APPROVED

Ankita
9/12/2019 12:16:06 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 12 02:56:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091119.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 12 02:33:14 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

