

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ091219.M
 Title : GC EXTRACTABLES
 Last Update : Fri Sep 13 03:48:31 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ043515.D 1000 =PQ043513.D 750 =PQ043514.D
 250 =PQ043516.D 50 =PQ043517.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.548	2.507	2.541	2.441	2.627	2.533 E6	2.67
2) SA	Decachlorobiphenyl	8.155	7.584	8.426	8.224	9.009	8.280 E6	6.21
3) L1	AR-1016-1	1.028	0.974	1.024	1.045	1.093	1.033 E5	4.14
4) L1	AR-1016-2	1.473	1.376	1.463	1.486	1.535	1.467 E5	3.93
5) L1	AR-1016-3	0.906	0.848	0.896	0.925	1.014	0.918 E5	6.60
6) L1	AR-1016-4	7.533	7.060	7.417	7.534	7.256	7.360 E4	2.75
7) L1	AR-1016-5	7.680	7.279	7.644	8.024	8.025	7.730 E4	4.02
8) L2	AR-1221-1	2.939					2.939 E4	0.00
9) L2	AR-1221-2	2.189					2.189 E4	0.00
10) L2	AR-1221-3	6.963					6.963 E4	0.00
11) L3	AR-1232-1	9.812					9.812 E4	0.00
12) L3	AR-1232-2	5.396					5.396 E4	0.00
13) L3	AR-1232-3	1.140					1.140 E5	0.00
14) L3	AR-1232-4	5.795					5.795 E4	0.00
15) L3	AR-1232-5	4.048					4.048 E4	0.00
16) L4	AR-1242-1	8.904					8.904 E4	0.00
17) L4	AR-1242-2	1.281					1.281 E5	0.00
18) L4	AR-1242-3	8.049					8.049 E4	0.00
19) L4	AR-1242-4	6.556					6.556 E4	0.00
20) L4	AR-1242-5	9.074					9.074 E4	0.00
21) L5	AR-1248-1	6.238					6.238 E4	0.00
22) L5	AR-1248-2	9.112					9.112 E4	0.00
23) L5	AR-1248-3	1.037					1.037 E5	0.00
24) L5	AR-1248-4	1.359					1.359 E5	0.00
25) L5	AR-1248-5	1.326					1.326 E5	0.00
26) L6	AR-1254-1	1.470					1.470 E5	0.00
27) L6	AR-1254-2	2.448					2.448 E5	0.00
28) L6	AR-1254-3	2.884					2.884 E5	0.00
29) L6	AR-1254-4	2.335					2.335 E5	0.00
30) L6	AR-1254-5	2.659					2.659 E5	0.00
31) L7	AR-1260-1	1.786	1.645	1.941	1.963	1.933	1.854 E5	7.33
32) L7	AR-1260-2	2.287	2.134	2.337	2.467	2.559	2.357 E5	6.98
33) L7	AR-1260-3	1.859	1.748	1.861	1.918	2.153	1.908 E5	7.88
34) L7	AR-1260-4	2.258	2.183	2.246	2.283	2.574	2.309 E5	6.62
35) L7	AR-1260-5	5.167	4.999	5.432	5.208	5.640	5.289 E5	4.72
36) L8	AR-1262-1	3.161					3.161 E5	0.00
37) L8	AR-1262-2	6.828					6.828 E5	0.00
38) L8	AR-1262-3	4.739					4.739 E5	0.00
39) L8	AR-1262-4	3.604					3.604 E5	0.00
40) L8	AR-1262-5	2.971					2.971 E5	0.00
41) L9	AR-1268-1	8.029					8.029 E5	0.00
42) L9	AR-1268-2	7.707					7.707 E5	0.00
43) L9	AR-1268-3	6.593					6.593 E5	0.00
44) L9	AR-1268-4	3.211					3.211 E5	0.00
45) L9	AR-1268-5	2.660					2.660 E6	0.00

Signal #2 Calibration Files

500 =PQ043515.D 1000 =PQ043513.D 750 =PQ043514.D
 250 =PQ043516.D 50 =PQ043517.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.023	1.005	1.015	1.003	1.098	1.029 E6	3.85
2) SA	Decachlorobiphenyl	3.702	3.363	3.771	3.783	4.186	3.761 E6	7.79
3) L1	AR-1016-1	4.410	4.159	4.433	4.484	5.090	4.515 E4	7.65
4) L1	AR-1016-2	5.956	5.484	5.820	5.918	6.123	5.860 E4	4.04
5) L1	AR-1016-3	3.154	2.947	3.137	3.152	3.594	3.197 E4	7.46
6) L1	AR-1016-4	2.623	2.455	2.593	2.718	3.223	2.722 E4	10.84

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ091219.M
 Title : GC EXTRACTABLES
 Last Update : Fri Sep 13 03:48:31 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ043515.D 1000 =PQ043513.D 750 =PQ043514.D
 250 =PQ043516.D 50 =PQ043517.D

Compound			500	1000	750	250	50	Avg	%RSD

7)	L1	AR-1016-5	3.447	3.280	3.426	3.594	3.906	3.530 E4	6.73
8)	L2	AR-1221-1	1.102					1.102 E4	0.00
9)	L2	AR-1221-2	8.654					8.654 E3	0.00
10)	L2	AR-1221-3	2.871					2.871 E4	0.00
11)	L3	AR-1232-1	4.077					4.077 E4	0.00
12)	L3	AR-1232-2	4.664					4.664 E4	0.00
13)	L3	AR-1232-3	2.503					2.503 E4	0.00
14)	L3	AR-1232-4	2.344					2.344 E4	0.00
15)	L3	AR-1232-5	2.566					2.566 E4	0.00
16)	L4	AR-1242-1	3.918					3.918 E4	0.00
17)	L4	AR-1242-2	5.022					5.022 E4	0.00
18)	L4	AR-1242-3	2.776					2.776 E4	0.00
19)	L4	AR-1242-4	2.881					2.881 E4	0.00
20)	L4	AR-1242-5	4.202					4.202 E4	0.00
21)	L5	AR-1248-1	2.674					2.674 E4	0.00
22)	L5	AR-1248-2	3.637					3.637 E4	0.00
23)	L5	AR-1248-3	3.791					3.791 E4	0.00
24)	L5	AR-1248-4	4.622					4.622 E4	0.00
25)	L5	AR-1248-5	5.163					5.163 E4	0.00
26)	L6	AR-1254-1	8.903					8.903 E4	0.00
27)	L6	AR-1254-2	8.131					8.131 E4	0.00
28)	L6	AR-1254-3	1.412					1.412 E5	0.00
29)	L6	AR-1254-4	9.751					9.751 E4	0.00
30)	L6	AR-1254-5	1.388					1.388 E5	0.00
31)	L7	AR-1260-1	7.880	7.340	7.661	8.259	9.615	8.151 E4	10.85
32)	L7	AR-1260-2	1.024	0.947	1.002	1.072	1.260	1.061 E5	11.28
33)	L7	AR-1260-3	0.934	0.882	0.916	0.975	1.082	0.958 E5	8.05
34)	L7	AR-1260-4	0.840	0.799	0.833	0.889	1.014	0.875 E5	9.62
35)	L7	AR-1260-5	2.152	2.082	2.180	2.177	2.369	2.192 E5	4.87
36)	L8	AR-1262-1	1.403					1.403 E5	0.00
37)	L8	AR-1262-2	2.595					2.595 E5	0.00
38)	L8	AR-1262-3	1.176					1.176 E5	0.00
39)	L8	AR-1262-4	2.373					2.373 E5	0.00
40)	L8	AR-1262-5	1.462					1.462 E5	0.00
41)	L9	AR-1268-1	3.244					3.244 E5	0.00
42)	L9	AR-1268-2	3.131					3.131 E5	0.00
43)	L9	AR-1268-3	2.827					2.827 E5	0.00
44)	L9	AR-1268-4	1.578					1.578 E5	0.00
45)	L9	AR-1268-5	1.218					1.218 E6	0.00

(#) = Out of Range