

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091421\
 Data File : PQ054457.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Sep 2021 00:31
 Operator : AJ\MA
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ091421AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 15 07:41:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 15 05:25:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.201	4.207	1267.3E6	682.2E6	53.066	51.607
2) SA Decachlor...	11.367	9.538	872.2E6	722.6E6	51.678	53.322
Target Compounds						
21) L5 AR-1248-1	6.525	5.467	233.2E6	127.9E6	465.844	488.897
22) L5 AR-1248-2	6.816	5.730	327.6E6	184.7E6	504.284	518.456
23) L5 AR-1248-3	7.036	5.775	370.7E6	187.2E6	518.371	526.441
24) L5 AR-1248-4	7.464	5.960	413.1E6	229.6E6	522.236	534.448
25) L5 AR-1248-5	7.506	6.383	409.8E6	229.5E6	510.652	535.683

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091421\
Data File : PQ054457.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Sep 2021 00:31
Operator : AJ\MA
Sample : AR1248ICV500
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ091421AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 15 07:41:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091421.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 15 05:25:53 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

