

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091619\
 Data File : PQ043598.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Sep 2019 19:25
 Operator : SM\AJ
 Sample : K3591-09
 Misc : AR1242 MDL 25 PPB
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 MDL-MDL-WATER-03-QT3-2019

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 04:50:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 13 03:48:31 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.702	4.787	22391379	9849041	8.840	9.572
2)	SA Decachlor...	12.095	10.559	110.0E6	42978007	13.285	11.427
Target Compounds							
3)	L1 AR-1016-1	7.170	6.242	2759803	404017	26.724	8.948 #
4)	L1 AR-1016-2	7.170	6.263	2759803	583105	18.817	9.950 #
5)	L1 AR-1016-3	7.239	6.530	1264820	163879	13.782	5.126 #
6)	L1 AR-1016-4	7.352	6.530	1007147	163879	13.684	6.020 #
7)	L1 AR-1016-5	7.688	6.785	2079921	236839	26.906	6.708 #
8)	L2 AR-1221-1	5.991	5.030	3170315	251764	107.859	22.846 #
9)	L2 AR-1221-2	6.091	5.204	1423668	248221	65.050	28.682 #
10)	L2 AR-1221-3	6.185	5.317	1401268	587612	20.126	20.464
11)	L3 AR-1232-1	6.185	5.317	1401268	587612	14.281	14.413
12)	L3 AR-1232-2	6.826	6.263	3043137	583105	56.399	12.501 #
13)	L3 AR-1232-3	7.170	6.530	2759803	163879	24.216	6.548 #
14)	L3 AR-1232-4	7.352	6.584	1007147	579721	17.381	24.730 #
15)	L3 AR-1232-5	7.461	6.785	661328	236839	16.337	9.230 #
16)	L4 AR-1242-1	7.170	6.242	2759803	404017	30.997	10.313 #
17)	L4 AR-1242-2	7.170	6.263	2759803	583105	21.543	11.611 #
18)	L4 AR-1242-3	7.239	6.530	1264820	163879	15.715	5.904 #
19)	L4 AR-1242-4	7.352	6.584	1007147	579721	15.363	20.123 #
20)	L4 AR-1242-5	8.175	7.193	5147162	1035387	56.722	24.643 #
21)	L5 AR-1248-1	7.170	6.242	2759803	404017	44.240	15.109 #
22)	L5 AR-1248-2	7.461	6.530	661328	163879	7.258	4.506 #
23)	L5 AR-1248-3	7.688	6.584	2079921	579721	20.049	15.290
24)	L5 AR-1248-4	8.134	6.785	3695764	236839	27.203	5.124 #
25)	L5 AR-1248-5	8.175	7.240	5147162	1820707	38.824	35.262
26)	L6 AR-1254-1	8.104	7.193	2972693	1035387	20.220	11.629 #
27)	L6 AR-1254-2	8.344	7.353	5552528	1925571	22.683	23.682
28)	L6 AR-1254-3	8.737	7.809	5145017	2090024	17.842	14.800
29)	L6 AR-1254-4	9.040	8.057	4689877	1306083	20.088	13.394 #
30)	L6 AR-1254-5	9.461	8.451	775499	1247095	2.916	8.988 #
31)	L7 AR-1260-1	8.938	7.943	3166722	2348967	17.083	28.818 #
32)	L7 AR-1260-2	9.181	8.139	2249183	1985443	9.543	18.712 #
33)	L7 AR-1260-3	9.567	8.334	982761	4468195	5.152	46.641 #
34)	L7 AR-1260-4	9.781	8.793	360595	978443	1.562	11.182 #
35)	L7 AR-1260-5	10.165	9.071	1339066	1327950	2.532	6.058 #
36)	L8 AR-1262-1	9.567	0.000	982761	0	3.109	N.D. #
37)	L8 AR-1262-2	10.165	9.071	1339066	1327950	1.961	5.117 #
38)	L8 AR-1262-3	10.535	0.000	1287644	0	2.717	N.D. #
39)	L8 AR-1262-4	10.565	9.458	479431	1010870	1.330	4.261 #
40)	L8 AR-1262-5	11.232	9.955	1458888	1824454	4.911	12.481 #
41)	L9 AR-1268-1	10.535	0.000	1287644	0	1.604	N.D. #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091619\
 Data File : PQ043598.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Sep 2019 19:25
 Operator : SM\AJ
 Sample : K3591-09
 Misc : AR1242 MDL 25 PPB
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 ECD_Q
 ClientSampleId :
 MDL-MDL-WATER-03-QT3-2019

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 04:50:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 13 03:48:31 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	10.565	9.458	479431	1010870	0.622	3.228 #
43) L9	AR-1268-3	10.822	9.649	1928531	506205	2.925	1.791 #
44) L9	AR-1268-4	11.232	9.955	1458888	1824454	4.544	11.559 #
45) L9	AR-1268-5	11.728	10.277	809462	1569698	0.304	1.289 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.



#1 Tetrachloro-m-xylene


R.T.: 5.702 min
Delta R.T.: 0.004 min
Response: 22391379
Conc: 8.84 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#1 Tetrachloro-m-xylene

R.T.: 4.787 min
Delta R.T.: 0.004 min
Response: 9849041
Conc: 9.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 12.095 min
Delta R.T.: 0.003 min
Response: 109999392
Conc: 13.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.559 min
Delta R.T.: 0.000 min
Response: 42978007
Conc: 11.43 ng/ml



#3 AR-1016-1


R.T.: 7.170 min
Delta R.T.: 0.028 min
Response: 2759803
Conc: 26.72 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#3 AR-1016-1

R.T.: 6.242 min
Delta R.T.: 0.003 min
Response: 404017
Conc: 8.95 ng/ml




#4 AR-1016-2

R.T.: 7.170 min
Delta R.T.: 0.002 min
Response: 2759803
Conc: 18.82 ng/ml



#4 AR-1016-2


R.T.: 6.263 min
Delta R.T.: 0.001 min
Response: 583105
Conc: 9.95 ng/ml



#5 AR-1016-3


R.T.: 7.239 min
Delta R.T.: 0.001 min
Response: 1264820
Conc: 13.78 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#5 AR-1016-3

R.T.: 6.530 min
Delta R.T.: 0.054 min
Response: 163879
Conc: 5.13 ng/ml




#6 AR-1016-4

R.T.: 7.352 min
Delta R.T.: 0.002 min
Response: 1007147
Conc: 13.68 ng/ml



#6 AR-1016-4


R.T.: 6.530 min
Delta R.T.: 0.000 min
Response: 163879
Conc: 6.02 ng/ml



#7 AR-1016-5


R.T.: 7.688 min
Delta R.T.: 0.003 min
Response: 2079921
Conc: 26.91 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#7 AR-1016-5

R.T.: 6.785 min
Delta R.T.: 0.002 min
Response: 236839
Conc: 6.71 ng/ml




#8 AR-1221-1

R.T.: 5.991 min
Delta R.T.: 0.013 min
Response: 3170315
Conc: 107.86 ng/ml



#8 AR-1221-1


R.T.: 5.030 min
Delta R.T.: -0.065 min
Response: 251764
Conc: 22.85 ng/ml



#9 AR-1221-2


R.T.: 6.091 min
Delta R.T.: 0.002 min
Response: 1423668
Conc: 65.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#9 AR-1221-2

R.T.: 5.204 min
Delta R.T.: -0.009 min
Response: 248221
Conc: 28.68 ng/ml




#10 AR-1221-3

R.T.: 6.185 min
Delta R.T.: 0.000 min
Response: 1401268
Conc: 20.13 ng/ml



#10 AR-1221-3


R.T.: 5.317 min
Delta R.T.: 0.000 min
Response: 587612
Conc: 20.46 ng/ml



#11 AR-1232-1


R.T.: 6.185 min
Delta R.T.: 0.000 min
Response: 1401268
Conc: 14.28 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#11 AR-1232-1

R.T.: 5.317 min
Delta R.T.: 0.000 min
Response: 587612
Conc: 14.41 ng/ml




#12 AR-1232-2

R.T.: 6.826 min
Delta R.T.: -0.002 min
Response: 3043137
Conc: 56.40 ng/ml



#12 AR-1232-2


R.T.: 6.263 min
Delta R.T.: -0.002 min
Response: 583105
Conc: 12.50 ng/ml



#13 AR-1232-3


R.T.: 7.170 min
Delta R.T.: 0.000 min
Response: 2759803
Conc: 24.22 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#13 AR-1232-3

R.T.: 6.530 min
Delta R.T.: 0.052 min
Response: 163879
Conc: 6.55 ng/ml




#14 AR-1232-4

R.T.: 7.352 min
Delta R.T.: -0.002 min
Response: 1007147
Conc: 17.38 ng/ml



#14 AR-1232-4


R.T.: 6.584 min
Delta R.T.: 0.002 min
Response: 579721
Conc: 24.73 ng/ml



#15 AR-1232-5


R.T.: 7.461 min
Delta R.T.: 0.000 min
Response: 661328
Conc: 16.34 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#15 AR-1232-5

R.T.: 6.785 min
Delta R.T.: 0.000 min
Response: 236839
Conc: 9.23 ng/ml




#16 AR-1242-1

R.T.: 7.170 min
Delta R.T.: 0.024 min
Response: 2759803
Conc: 31.00 ng/ml



#16 AR-1242-1


R.T.: 6.242 min
Delta R.T.: 0.000 min
Response: 404017
Conc: 10.31 ng/ml



#17 AR-1242-2


R.T.: 7.170 min
Delta R.T.: -0.001 min
Response: 2759803
Conc: 21.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#17 AR-1242-2

R.T.: 6.263 min
Delta R.T.: -0.003 min
Response: 583105
Conc: 11.61 ng/ml




#18 AR-1242-3

R.T.: 7.239 min
Delta R.T.: -0.003 min
Response: 1264820
Conc: 15.71 ng/ml



#18 AR-1242-3


R.T.: 6.530 min
Delta R.T.: 0.051 min
Response: 163879
Conc: 5.90 ng/ml



#19 AR-1242-4


R.T.: 7.352 min
Delta R.T.: -0.003 min
Response: 1007147
Conc: 15.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#19 AR-1242-4

R.T.: 6.584 min
Delta R.T.: 0.000 min
Response: 579721
Conc: 20.12 ng/ml




#20 AR-1242-5

R.T.: 8.175 min
Delta R.T.: -0.001 min
Response: 5147162
Conc: 56.72 ng/ml



#20 AR-1242-5


R.T.: 7.193 min
Delta R.T.: -0.002 min
Response: 1035387
Conc: 24.64 ng/ml



#21 AR-1248-1


R.T.: 7.170 min
Delta R.T.: 0.027 min
Response: 2759803
Conc: 44.24 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#21 AR-1248-1

R.T.: 6.242 min
Delta R.T.: 0.001 min
Response: 404017
Conc: 15.11 ng/ml




#22 AR-1248-2

R.T.: 7.461 min
Delta R.T.: 0.003 min
Response: 661328
Conc: 7.26 ng/ml



#22 AR-1248-2


R.T.: 6.530 min
Delta R.T.: -0.002 min
Response: 163879
Conc: 4.51 ng/ml



#23 AR-1248-3


R.T.: 7.688 min
Delta R.T.: 0.003 min
Response: 2079921
Conc: 20.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#23 AR-1248-3

R.T.: 6.584 min
Delta R.T.: 0.003 min
Response: 579721
Conc: 15.29 ng/ml




#24 AR-1248-4

R.T.: 8.134 min
Delta R.T.: 0.002 min
Response: 3695764
Conc: 27.20 ng/ml



#24 AR-1248-4


R.T.: 6.785 min
Delta R.T.: 0.000 min
Response: 236839
Conc: 5.12 ng/ml



#25 AR-1248-5


R.T.: 8.175 min
Delta R.T.: 0.002 min
Response: 5147162
Conc: 38.82 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#25 AR-1248-5

R.T.: 7.240 min
Delta R.T.: 0.000 min
Response: 1820707
Conc: 35.26 ng/ml




#26 AR-1254-1

R.T.: 8.104 min
Delta R.T.: 0.004 min
Response: 2972693
Conc: 20.22 ng/ml



#26 AR-1254-1


R.T.: 7.193 min
Delta R.T.: 0.001 min
Response: 1035387
Conc: 11.63 ng/ml



#27 AR-1254-2


R.T.: 8.344 min
Delta R.T.: 0.005 min
Response: 5552528
Conc: 22.68 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#27 AR-1254-2

R.T.: 7.353 min
Delta R.T.: -0.007 min
Response: 1925571
Conc: 23.68 ng/ml




#28 AR-1254-3

R.T.: 8.737 min
Delta R.T.: 0.005 min
Response: 5145017
Conc: 17.84 ng/ml



#28 AR-1254-3


R.T.: 7.809 min
Delta R.T.: 0.002 min
Response: 2090024
Conc: 14.80 ng/ml



#29 AR-1254-4


R.T.: 9.040 min
Delta R.T.: 0.005 min
Response: 4689877
Conc: 20.09 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#29 AR-1254-4

R.T.: 8.057 min
Delta R.T.: 0.001 min
Response: 1306083
Conc: 13.39 ng/ml




#30 AR-1254-5

R.T.: 9.461 min
Delta R.T.: -0.011 min
Response: 775499
Conc: 2.92 ng/ml



#30 AR-1254-5


R.T.: 8.451 min
Delta R.T.: -0.051 min
Response: 1247095
Conc: 8.99 ng/ml



#31 AR-1260-1


R.T.: 8.938 min
Delta R.T.: 0.031 min
Response: 3166722
Conc: 17.08 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#31 AR-1260-1

R.T.: 7.943 min
Delta R.T.: -0.008 min
Response: 2348967
Conc: 28.82 ng/ml




#32 AR-1260-2

R.T.: 9.181 min
Delta R.T.: 0.003 min
Response: 2249183
Conc: 9.54 ng/ml



#32 AR-1260-2


R.T.: 8.139 min
Delta R.T.: -0.013 min
Response: 1985443
Conc: 18.71 ng/ml



#33 AR-1260-3


R.T.: 9.567 min
Delta R.T.: 0.015 min
Response: 982761
Conc: 5.15 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#33 AR-1260-3

R.T.: 8.334 min
Delta R.T.: 0.015 min
Response: 4468195
Conc: 46.64 ng/ml




#34 AR-1260-4

R.T.: 9.781 min
Delta R.T.: -0.013 min
Response: 360595
Conc: 1.56 ng/ml



#34 AR-1260-4


R.T.: 8.793 min
Delta R.T.: -0.025 min
Response: 978443
Conc: 11.18 ng/ml



#35 AR-1260-5


R.T.: 10.165 min
Delta R.T.: 0.025 min
Response: 1339066
Conc: 2.53 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#35 AR-1260-5

R.T.: 9.071 min
Delta R.T.: 0.003 min
Response: 1327950
Conc: 6.06 ng/ml




#36 AR-1262-1

R.T.: 9.567 min
Delta R.T.: 0.019 min
Response: 982761
Conc: 3.11 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 8.542 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 10.165 min
Delta R.T.: 0.029 min
Response: 1339066
Conc: 1.96 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#37 AR-1262-2

R.T.: 9.071 min
Delta R.T.: 0.005 min
Response: 1327950
Conc: 5.12 ng/ml




#38 AR-1262-3

R.T.: 10.535 min
Delta R.T.: 0.064 min
Response: 1287644
Conc: 2.72 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T.: 9.361 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 10.565 min
Delta R.T.: -0.003 min
Response: 479431
Conc: 1.33 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#39 AR-1262-4

R.T.: 9.458 min
Delta R.T.: 0.031 min
Response: 1010870
Conc: 4.26 ng/ml




#40 AR-1262-5

R.T.: 11.232 min
Delta R.T.: -0.046 min
Response: 1458888
Conc: 4.91 ng/ml



#40 AR-1262-5


R.T.: 9.955 min
Delta R.T.: 0.002 min
Response: 1824454
Conc: 12.48 ng/ml



#41 AR-1268-1


R.T.: 10.535 min
Delta R.T.: 0.064 min
Response: 1287644
Conc: 1.60 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 9.362 min
Response: 0
Conc: N.D.




#42 AR-1268-2

R.T.: 10.565 min
Delta R.T.: -0.009 min
Response: 479431
Conc: 0.62 ng/ml



#42 AR-1268-2


R.T.: 9.458 min
Delta R.T.: 0.029 min
Response: 1010870
Conc: 3.23 ng/ml



#43 AR-1268-3


R.T.: 10.822 min
Delta R.T.: 0.004 min
Response: 1928531
Conc: 2.92 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019




#43 AR-1268-3

R.T.: 9.649 min
Delta R.T.: 0.003 min
Response: 506205
Conc: 1.79 ng/ml




#44 AR-1268-4

R.T.: 11.232 min
Delta R.T.: -0.046 min
Response: 1458888
Conc: 4.54 ng/ml



#44 AR-1268-4


R.T.: 9.955 min
Delta R.T.: 0.000 min
Response: 1824454
Conc: 11.56 ng/ml



#45 AR-1268-5

R.T.: 11.728 min
Delta R.T.: 0.004 min
Response: 809462
Conc: 0.30 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-MDL-WATER-03-QT3-2019



#45 AR-1268-5

R.T.: 10.277 min
Delta R.T.: 0.002 min
Response: 1569698
Conc: 1.29 ng/ml