

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091624\
 Data File : PQ068630.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Sep 2024 10:57
 Operator : YP\AJ
 Sample : P3991-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 A0BD0

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/17/2024
 Supervised By :Ankita Jodhani 09/17/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 02:23:04 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082724CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 28 02:08:25 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	3.437	2.820	68326760	64570001	8.648m	9.117
2) SA Decachlor...	8.609	7.597	67584503	101.0E6	29.715	29.853

Target Compounds

31) L7 AR-1260-1	6.094	5.244	8944.1E6	12444.5E6	27103.238	34216.198 #
32) L7 AR-1260-2	6.353	5.433	12866.8E6	15639.8E6	34146.962	35632.721
33) L7 AR-1260-3	6.706	5.575	11033.3E6	11395.0E6	39900.042	26437.351 #
34) L7 AR-1260-4	6.924	6.037	12279.3E6	12633.9E6	40218.118	37203.471
35) L7 AR-1260-5	7.234	6.286	27686.1E6	29753.3E6	47438.453	37489.105

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091624\
Data File : PQ068630.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Sep 2024 10:57
Operator : YP\AJ
Sample : P3991-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
A0BD0

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/17/2024
Supervised By :Ankita Jodhani 09/17/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 17 02:23:04 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082724CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 28 02:08:25 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

