

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ091719CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 18 04:28:29 2019
 Response Via : Initial Calibration

Calibration Files

400 =PQ043609.D 1600 =PQ043611.D 800 =PQ043610.D
 200 =PQ043608.D 100 =PQ043607.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.178	2.045	2.252	2.203	2.018	2.139	E6	4.78
2) SA	Decachlorobiphenyl	7.772	7.222	7.852	8.119	8.221	7.837	E6	4.98
3) L1	AR-1016-1	0.960	0.893	1.018	1.029	0.989	0.978	E5	5.55
4) L1	AR-1016-2	1.398	1.319	1.490	1.486	1.413	1.421	E5	4.96
5) L1	AR-1016-3	8.451	7.905	8.977	9.017	8.702	8.610	E4	5.29
6) L1	AR-1016-4	7.138	6.677	7.600	7.698	7.473	7.317	E4	5.68
7) L1	AR-1016-5	7.650	6.964	8.102	8.119	7.916	7.750	E4	6.17
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1	1.836	1.644	1.884	2.012	2.140	1.903	E5	9.84
32) L7	AR-1260-2	2.383	2.167	2.465	2.582	2.624	2.444	E5	7.44
33) L7	AR-1260-3	1.986	1.809	2.035	2.085	2.054	1.994	E5	5.49
34) L7	AR-1260-4	2.495	2.317	2.584	2.591	2.572	2.512	E5	4.59
35) L7	AR-1260-5	5.615	5.379	5.875	5.711	5.727	5.661	E5	3.24
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ043609.D 1600 =PQ043611.D 800 =PQ043610.D
 200 =PQ043608.D 100 =PQ043607.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	9.031	8.668	9.463	9.530	9.140	9.167	E5	3.81
2) SA	Decachlorobiphenyl	3.051	2.810	3.061	3.152	3.203	3.055	E6	4.95
3) L1	AR-1016-1	3.725	3.435	3.939	4.040	3.833	3.794	E4	6.13
4) L1	AR-1016-2	5.128	4.753	5.401	5.556	5.271	5.222	E4	5.87
5) L1	AR-1016-3	2.740	2.601	2.940	2.985	2.796	2.812	E4	5.51
6) L1	AR-1016-4	2.248	2.086	2.385	2.541	2.472	2.346	E4	7.75

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ091719CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 18 04:28:29 2019
 Response Via : Initial Calibration

Calibration Files

400 =PQ043609.D 1600 =PQ043611.D 800 =PQ043610.D
 200 =PQ043608.D 100 =PQ043607.D

Compound			400	1600	800	200	100	Avg	%RSD
7)	L1	AR-1016-5	3.116	2.811	3.191	3.697	3.187	3.200 E4	9.95
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1	6.898	6.165	7.143	7.410	7.350	6.993 E4	7.21
32)	L7	AR-1260-2	8.938	7.834	9.187	9.638	9.466	9.013 E4	7.89
33)	L7	AR-1260-3	8.488	7.713	8.808	9.104	9.049	8.632 E4	6.59
34)	L7	AR-1260-4	7.721	7.033	7.944	8.059	8.179	7.787 E4	5.83
35)	L7	AR-1260-5	2.071	1.989	2.177	2.152	2.164	2.111 E5	3.77
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

(#) = Out of Range