

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ091719CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 18 04:46:26 2019
 Response Via : Initial Calibration

Calibration Files

400 =PQ043614.D 1600 =PQ043616.D 800 =PQ043615.D
 200 =PQ043613.D 100 =PQ043612.D

	Compound	400	1600	800	200	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.233	2.049	2.112	2.209	2.316	2.184	E6	4.80
2) SA	Decachlorobiphenyl	0.848	0.774	0.837	1.028	0.869	0.871	E7	10.84
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1	2.443	2.109	2.200	2.331	2.360	2.289	E4	5.82
9) L2	AR-1221-2	1.729	1.522	1.532	1.844	2.022	1.730	E4	12.28
10) L2	AR-1221-3	6.046	5.377	5.666	6.163	6.879	6.026	E4	9.45
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ043614.D 1600 =PQ043616.D 800 =PQ043615.D
 200 =PQ043613.D 100 =PQ043612.D

	Compound	400	1600	800	200	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	9.160	8.517	8.657	9.185	9.580	9.020	E5	4.79
2) SA	Decachlorobiphenyl	3.226	2.958	3.190	3.762	3.428	3.313	E6	9.10
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ091719CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 18 04:46:26 2019
 Response Via : Initial Calibration

Calibration Files

400 =PQ043614.D 1600 =PQ043616.D 800 =PQ043615.D
 200 =PQ043613.D 100 =PQ043612.D

Compound			400	1600	800	200	100	Avg	%RSD
7)	L1	AR-1016-5						0.000	-1.00
8)	L2	AR-1221-1	1.145	0.972	1.021	1.114	1.208	1.092 E4	8.70
9)	L2	AR-1221-2	7.912	6.819	7.235	8.453	9.340	7.952 E3	12.54
10)	L2	AR-1221-3	2.808	2.397	2.562	2.918	3.005	2.738 E4	9.24
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1						0.000	-1.00
32)	L7	AR-1260-2						0.000	-1.00
33)	L7	AR-1260-3						0.000	-1.00
34)	L7	AR-1260-4						0.000	-1.00
35)	L7	AR-1260-5						0.000	-1.00
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

(#) = Out of Range