

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ091719CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 18 05:40:14 2019
 Response Via : Initial Calibration

Calibration Files

400 =PQ043629.D 1600 =PQ043631.D 800 =PQ043630.D
 200 =PQ043628.D 100 =PQ043627.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.693	2.436	2.133	2.583	2.936	2.556	E6	11.69
2) SA	Decachlorobiphenyl	0.918	0.843	0.880	0.977	1.028	0.929	E7	8.00
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	7.109	6.390	6.080	7.246	8.414	7.048	E4	12.85
22) L5	AR-1248-2	1.022	0.907	0.878	1.054	1.219	1.016	E5	13.32
23) L5	AR-1248-3	1.215	1.096	1.046	1.232	1.434	1.205	E5	12.49
24) L5	AR-1248-4	1.593	1.442	1.401	1.690	1.873	1.600	E5	12.00
25) L5	AR-1248-5	1.552	1.401	1.370	1.687	1.796	1.561	E5	11.68
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ043629.D 1600 =PQ043631.D 800 =PQ043630.D
 200 =PQ043628.D 100 =PQ043627.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.231	1.134	1.011	1.117	1.378	1.174	E6	11.78
2) SA	Decachlorobiphenyl	3.499	3.220	3.326	3.735	3.912	3.539	E6	8.07
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ091719CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 18 05:40:14 2019
 Response Via : Initial Calibration

Calibration Files

400 =PQ043629.D 1600 =PQ043631.D 800 =PQ043630.D
 200 =PQ043628.D 100 =PQ043627.D

Compound			400	1600	800	200	100	Avg	%RSD

7) L1	AR-1016-5							0.000	-1.00
8) L2	AR-1221-1							0.000	-1.00
9) L2	AR-1221-2							0.000	-1.00
10) L2	AR-1221-3							0.000	-1.00
11) L3	AR-1232-1							0.000	-1.00
12) L3	AR-1232-2							0.000	-1.00
13) L3	AR-1232-3							0.000	-1.00
14) L3	AR-1232-4							0.000	-1.00
15) L3	AR-1232-5							0.000	-1.00
16) L4	AR-1242-1							0.000	-1.00
17) L4	AR-1242-2							0.000	-1.00
18) L4	AR-1242-3							0.000	-1.00
19) L4	AR-1242-4							0.000	-1.00
20) L4	AR-1242-5							0.000	-1.00
21) L5	AR-1248-1		2.781	2.473	2.371	2.861	3.311	2.759 E4	13.42
22) L5	AR-1248-2		3.820	3.284	3.171	3.962	4.585	3.764 E4	15.14
23) L5	AR-1248-3		3.908	3.408	3.273	4.008	4.637	3.847 E4	14.10
24) L5	AR-1248-4		4.728	4.284	4.076	4.869	5.577	4.707 E4	12.39
25) L5	AR-1248-5		5.285	4.749	4.601	5.489	6.202	5.265 E4	12.14
26) L6	AR-1254-1							0.000	-1.00
27) L6	AR-1254-2							0.000	-1.00
28) L6	AR-1254-3							0.000	-1.00
29) L6	AR-1254-4							0.000	-1.00
30) L6	AR-1254-5							0.000	-1.00
31) L7	AR-1260-1							0.000	-1.00
32) L7	AR-1260-2							0.000	-1.00
33) L7	AR-1260-3							0.000	-1.00
34) L7	AR-1260-4							0.000	-1.00
35) L7	AR-1260-5							0.000	-1.00
36) L8	AR-1262-1							0.000	-1.00
37) L8	AR-1262-2							0.000	-1.00
38) L8	AR-1262-3							0.000	-1.00
39) L8	AR-1262-4							0.000	-1.00
40) L8	AR-1262-5							0.000	-1.00
41) L9	AR-1268-1							0.000	-1.00
42) L9	AR-1268-2							0.000	-1.00
43) L9	AR-1268-3							0.000	-1.00
44) L9	AR-1268-4							0.000	-1.00
45) L9	AR-1268-5							0.000	-1.00

(#) = Out of Range