

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091818\
 Data File : PQ032010.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2018 22:14
 Operator : SM\SJ
 Sample : J4947-01
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 97-20

Manual Integrations
 APPROVED

Sohil
 9/19/2018 2:16:32 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 05:03:46 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 19 02:07:33 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.598	3.880	20099537	25036275	8.668m	12.501 #
2) SA Decachlor...	10.451	8.951	26411607	23604313	12.670	12.190m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091818\
Data File : PQ032010.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2018 22:14
Operator : SM\SJ
Sample : J4947-01
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
97-20


Manual Integrations
APPROVED

Sohil
9/19/2018 2:16:32 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 05:03:46 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091818.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 19 02:07:33 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.598 min
Delta R.T.: 0.004 min
Response: 20099537
Conc: 8.67 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
97-20


Manual Integrations
APPROVED

Sohil
9/19/2018 2:16:32 PM




#1 Tetrachloro-m-xylene

R.T.: 3.880 min
Delta R.T.: 0.007 min
Response: 25036275
Conc: 12.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.451 min
Delta R.T.: 0.000 min
Response: 26411607
Conc: 12.67 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.951 min
Delta R.T.: 0.000 min
Response: 23604313
Conc: 12.19 ng/ml m