

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091922\
 Data File : PQ059136.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2022 20:45
 Operator : YP\AJ
 Sample : AR1254ICC100
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12541046

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 05:50:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091922CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 05:47:55 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.672	4.010	64585473	42505603	4.968	4.933
2) SA Decachlor...	10.558	9.099	59506629	49146569	10.607	10.869
Target Compounds						
26) L6 AR-1254-1	6.708	5.908	44220636	47465545	107.580	109.152
27) L6 AR-1254-2	6.925	6.053	65408383	43521877	105.763	114.689
28) L6 AR-1254-3	7.296	6.461	65870873	63609744	104.267	106.864
29) L6 AR-1254-4	7.582	6.688	46004401	41793056	105.252	98.512
30) L6 AR-1254-5	8.000	7.105	47308112	52895329	104.314	111.188

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091922\
Data File : PQ059136.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2022 20:45
Operator : YP\AJ
Sample : AR1254ICC100
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12541046

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 05:50:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091922CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 05:47:55 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

