

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091922\
 Data File : PQ059166.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2022 04:02
 Operator : YP\AJ
 Sample : N4559-09
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 A0AY4

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 06:59:36 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091922CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 06:48:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.671	4.008	220.8E6	139.7E6	18.904	18.214
2) SA Decachlor...	10.558	9.098	150.5E6	119.8E6	29.162	28.759
Target Compounds						
31) L7 AR-1260-1	7.454	6.587	2604.4E6	1865.9E6	6789.185	5692.646
32) L7 AR-1260-2	7.711	6.772	3087.2E6	2582.2E6	6878.560	6640.072
33) L7 AR-1260-3	8.069	6.930	1926.3E6	1987.0E6	6168.762	5640.340
34) L7 AR-1260-4	8.307	7.400	2406.0E6	1541.5E6	6891.510	5555.476
35) L7 AR-1260-5	8.647	7.638	4257.5E6	3881.1E6	6166.461	6092.375

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091922\
Data File : PQ059166.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2022 04:02
Operator : YP\AJ
Sample : N4559-09
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
A0AY4

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 06:59:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091922CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 06:48:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

