

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091923\
 Data File : PQ063281.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2023 00:32
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 04:42:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 20 04:41:30 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.463	2.781	495.9E6	375.7E6	96.835	96.528
2) SA Decachlor...	8.696	7.583	364.4E6	351.1E6	95.064	94.695
Target Compounds						
21) L5 AR-1248-1	4.578	3.794	96002037	76175487	938.572	933.465
22) L5 AR-1248-2	4.847	4.020	136.3E6	111.5E6	939.519	915.267
23) L5 AR-1248-3	5.040	4.056	164.7E6	108.1E6	938.782	926.606
24) L5 AR-1248-4	5.436	4.215	184.1E6	133.2E6	951.160	933.686
25) L5 AR-1248-5	5.471	4.592	181.7E6	133.4E6	962.137	947.633



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ091923\
Data File : PQ063281.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2023 00:32
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 04:42:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091923.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 20 04:41:30 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.463 min
Delta R.T.: 0.000 min
Response: 495891927
Conc: 96.83 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.781 min
Delta R.T.: 0.000 min
Response: 375705927
Conc: 96.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 364433037
Conc: 95.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.583 min
Delta R.T.: 0.000 min
Response: 351114417
Conc: 94.69 ng/ml



#21 AR-1248-1


R.T.: 4.578 min
Delta R.T.: 0.000 min
Response: 96002037
Conc: 938.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#21 AR-1248-1

R.T.: 3.794 min
Delta R.T.: 0.000 min
Response: 76175487
Conc: 933.47 ng/ml




#22 AR-1248-2

R.T.: 4.847 min
Delta R.T.: 0.000 min
Response: 136310529
Conc: 939.52 ng/ml



#22 AR-1248-2


R.T.: 4.020 min
Delta R.T.: 0.000 min
Response: 111496405
Conc: 915.27 ng/ml



#23 AR-1248-3


R.T.: 5.040 min
Delta R.T.: 0.001 min
Response: 164695444
Conc: 938.78 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#23 AR-1248-3

R.T.: 4.056 min
Delta R.T.: 0.000 min
Response: 108131007
Conc: 926.61 ng/ml




#24 AR-1248-4

R.T.: 5.436 min
Delta R.T.: 0.000 min
Response: 184104966
Conc: 951.16 ng/ml



#24 AR-1248-4


R.T.: 4.215 min
Delta R.T.: 0.000 min
Response: 133205764
Conc: 933.69 ng/ml



#25 AR-1248-5

R.T.: 5.471 min
Delta R.T.: 0.000 min
Response: 181716940
Conc: 962.14 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#25 AR-1248-5

R.T.: 4.592 min
Delta R.T.: 0.000 min
Response: 133449401
Conc: 947.63 ng/ml