

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092018\
 Data File : PQ032078.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2018 23:54
 Operator : SM\SJ
 Sample : AR1254ICC1600
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 03:58:16 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092018CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 21 03:54:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.596	3.875	156.5E6	138.3E6	77.999	78.352
2) SA Decachlor...	10.453	8.951	308.7E6	280.9E6	154.447	156.285
Target Compounds						
26) L6 AR-1254-1	6.631	5.780	129.0E6	163.0E6	1454.596	1488.156
27) L6 AR-1254-2	6.849	5.925	205.1E6	141.1E6	1462.995	1483.897
28) L6 AR-1254-3	7.218	6.333	230.1E6	248.3E6	1493.010	1504.781
29) L6 AR-1254-4	7.504	6.560	168.8E6	157.4E6	1491.120	1492.557
30) L6 AR-1254-5	7.923	6.979	187.3E6	217.1E6	1501.472	1522.888

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092018\
Data File : PQ032078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2018 23:54
Operator : SM\SJ
Sample : AR1254ICC1600
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 03:58:16 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092018CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 21 03:54:31 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

