

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092018\
 Data File : PQ032102.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2018 05:41
 Operator : SM\SJ
 Sample : J4860-01
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 A41S8

Manual Integrations
 APPROVED

mohammad
 9/25/2018 8:55:17 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 05:51:40 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092018CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 21 05:45:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.577	3.875	237.7E6	3428883	112.407m	1.963 #
2) SA Decachlor...	10.452	8.956	2786929	1214353	1.325m	0.655m#
Target Compounds						
11) L3 AR-1232-1	4.971	4.255	1000.7E6	832.4E6	28115.747m	26947.455m
12) L3 AR-1232-2	5.505	4.992	555.9E6	331.5E6	30071.986m	10022.195 #
13) L3 AR-1232-3	5.797	5.171	467.2E6	54358810	12152.289	3250.214m#
14) L3 AR-1232-4	5.959	5.251	58571597	207.9E6	3232.576m	13957.745m#
15) L3 AR-1232-5	6.047	5.428	180.3E6	77381801	12860.340m	4838.153 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092018\
Data File : PQ032102.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2018 05:41
Operator : SM\SJ
Sample : J4860-01
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
A41S8

Manual Integrations
APPROVED

mohammad
9/25/2018 8:55:17 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 05:51:40 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092018CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 21 05:45:25 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

