

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092022\
 Data File : PQ059196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2022 12:17
 Operator : YP\AJ
 Sample : N4691-13
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 PCB-GCP2-BLANK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 15:35:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091922CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 06:48:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.671	0.000	163973	0	0.014	N.D. #
2)	SA Decachlor...	10.563	9.117	164917	215525	0.032	0.052 #
Target Compounds							
3)	L1 AR-1016-1	5.847	5.100	619164	139142	1.870	0.659 #
4)	L1 AR-1016-2	5.873	5.100	289984	139142	0.586	0.372 #
5)	L1 AR-1016-3	5.930	5.303	111925	144880	0.364	0.873 #
6)	L1 AR-1016-4	6.056	5.342	507594	730387	2.032	5.434 #
7)	L1 AR-1016-5	6.335	5.541	132697	914532	0.602	5.279 #
8)	L2 AR-1221-1	4.887	4.230	383520	881211	3.212	11.513 #
9)	L2 AR-1221-2	4.974	4.295	414790	51814	4.938	0.983 #
10)	L2 AR-1221-3	5.051	4.385	208168	367155	0.725	2.022 #
11)	L3 AR-1232-1	5.051	4.385	208168	367155	0.899	2.531 #
12)	L3 AR-1232-2	5.583	5.100	235868	139142	2.114	0.896 #
13)	L3 AR-1232-3	5.873	5.303	289984	144880	1.372	2.182 #
14)	L3 AR-1232-4	6.056	5.375	507594	1508282	5.029	22.746 #
15)	L3 AR-1232-5	6.110	5.541	221802	914532	2.667	13.216 #
16)	L4 AR-1242-1	5.847	5.100	619164	139142	2.031	0.774 #
17)	L4 AR-1242-2	5.873	5.100	289984	139142	0.633	0.428 #
18)	L4 AR-1242-3	5.930	5.303	111925	144880	0.406	1.017 #
19)	L4 AR-1242-4	6.056	5.375	507594	1508282	2.282	9.927 #
20)	L4 AR-1242-5	6.775	5.919	128009	1968552	0.575	10.730 #
21)	L5 AR-1248-1	5.847	5.100	619164	139142	2.737	0.909 #
22)	L5 AR-1248-2	6.110	5.342	221802	730387	0.666	3.212 #
23)	L5 AR-1248-3	6.335	5.375	132697	1508282	0.370	6.037 #
24)	L5 AR-1248-4	6.734	5.541	341920	914532	0.877	3.270 #
25)	L5 AR-1248-5	6.775	5.956	128009	543200	0.342	1.929 #
26)	L6 AR-1254-1	6.705	5.919	289217	1968552	0.711	4.574 #
27)	L6 AR-1254-2	6.960f	6.067	307734	1323130	0.505	3.494 #
28)	L6 AR-1254-3	7.286	6.463	144669	1851940	0.232	3.148 #
29)	L6 AR-1254-4	7.580	6.669	346859	778757	0.804	1.899 #
30)	L6 AR-1254-5	7.998	7.104	113863	1194854	0.255	2.523 #
31)	L7 AR-1260-1	7.459	6.569	697368	1550852	1.818	4.731 #
32)	L7 AR-1260-2	7.708	6.783	404236	464594	0.901	1.195 #
33)	L7 AR-1260-3	8.096	6.924	194078	956241	0.622	2.714 #
34)	L7 AR-1260-4	8.318	7.387	122494	306757	0.351	1.106 #
35)	L7 AR-1260-5	8.652	7.641	324406	773840	0.470	1.215 #
36)	L8 AR-1262-1	8.096	7.127	194078	1273248	0.367	2.573 #
37)	L8 AR-1262-2	8.652	7.641	324406	773840	0.363	0.958 #
38)	L8 AR-1262-3	8.991	7.928	148735	539174	0.396	1.771 #
39)	L8 AR-1262-4	9.067	7.989	62586	331464	0.274	0.593 #
40)	L8 AR-1262-5	9.772	8.508	149017	811987	0.522	3.437 #
41)	L9 AR-1268-1	8.991	7.928	148735	539174	0.142	0.591 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092022\
 Data File : PQ059196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2022 12:17
 Operator : YP\AJ
 Sample : N4691-13
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 PCB-GCP2-BLANK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 15:35:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091922CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 06:48:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.075	7.989	45980	331464	0.049	0.402 #
43) L9	AR-1268-3	9.337	8.193	113087	1041223	0.144	1.501 #
44) L9	AR-1268-4	9.772	8.508	149017	811987	0.463	2.982 #
45) L9	AR-1268-5	10.248f	8.850	104204	285168	0.044	0.145 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092022\
 Data File : PQ059196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2022 12:17
 Operator : YP\AJ
 Sample : N4691-13
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 PCB-GCP2-BLANK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 15:35:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091922CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 06:48:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.671 min
Delta R.T.: -0.002 min
Response: 163973
Conc: 0.01 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 4.009 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 10.563 min
Delta R.T.: 0.000 min
Response: 164917
Conc: 0.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.117 min
Delta R.T.: 0.017 min
Response: 215525
Conc: 0.05 ng/ml



#3 AR-1016-1


R.T.: 5.847 min
Delta R.T.: -0.006 min
Response: 619164
Conc: 1.87 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#3 AR-1016-1

R.T.: 5.100 min
Delta R.T.: -0.002 min
Response: 139142
Conc: 0.66 ng/ml




#4 AR-1016-2

R.T.: 5.873 min
Delta R.T.: -0.004 min
Response: 289984
Conc: 0.59 ng/ml



#4 AR-1016-2


R.T.: 5.100 min
Delta R.T.: -0.021 min
Response: 139142
Conc: 0.37 ng/ml



#5 AR-1016-3


R.T.: 5.930 min
Delta R.T.: -0.009 min
Response: 111925
Conc: 0.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#5 AR-1016-3

R.T.: 5.303 min
Delta R.T.: 0.002 min
Response: 144880
Conc: 0.87 ng/ml




#6 AR-1016-4

R.T.: 6.056 min
Delta R.T.: 0.016 min
Response: 507594
Conc: 2.03 ng/ml



#6 AR-1016-4


R.T.: 5.342 min
Delta R.T.: 0.004 min
Response: 730387
Conc: 5.43 ng/ml



#7 AR-1016-5


R.T.: 6.335 min
Delta R.T.: 0.002 min
Response: 132697
Conc: 0.60 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#7 AR-1016-5

R.T.: 5.541 min
Delta R.T.: -0.014 min
Response: 914532
Conc: 5.28 ng/ml




#8 AR-1221-1

R.T.: 4.887 min
Delta R.T.: 0.000 min
Response: 383520
Conc: 3.21 ng/ml



#8 AR-1221-1


R.T.: 4.230 min
Delta R.T.: 0.008 min
Response: 881211
Conc: 11.51 ng/ml



#9 AR-1221-2


R.T.: 4.974 min
Delta R.T.: 0.000 min
Response: 414790
Conc: 4.94 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#9 AR-1221-2

R.T.: 4.295 min
Delta R.T.: -0.015 min
Response: 51814
Conc: 0.98 ng/ml




#10 AR-1221-3

R.T.: 5.051 min
Delta R.T.: -0.001 min
Response: 208168
Conc: 0.72 ng/ml



#10 AR-1221-3


R.T.: 4.385 min
Delta R.T.: -0.002 min
Response: 367155
Conc: 2.02 ng/ml



#11 AR-1232-1


R.T.: 5.051 min
Delta R.T.: -0.001 min
Response: 208168
Conc: 0.90 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#11 AR-1232-1

R.T.: 4.385 min
Delta R.T.: -0.002 min
Response: 367155
Conc: 2.53 ng/ml




#12 AR-1232-2

R.T.: 5.583 min
Delta R.T.: -0.001 min
Response: 235868
Conc: 2.11 ng/ml



#12 AR-1232-2


R.T.: 5.100 min
Delta R.T.: -0.022 min
Response: 139142
Conc: 0.90 ng/ml



#13 AR-1232-3


R.T.: 5.873 min
Delta R.T.: -0.003 min
Response: 289984
Conc: 1.37 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#13 AR-1232-3

R.T.: 5.303 min
Delta R.T.: 0.002 min
Response: 144880
Conc: 2.18 ng/ml




#14 AR-1232-4

R.T.: 6.056 min
Delta R.T.: 0.016 min
Response: 507594
Conc: 5.03 ng/ml



#14 AR-1232-4


R.T.: 5.375 min
Delta R.T.: -0.006 min
Response: 1508282
Conc: 22.75 ng/ml



#15 AR-1232-5


R.T.: 6.110 min
Delta R.T.: -0.014 min
Response: 221802
Conc: 2.67 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#15 AR-1232-5

R.T.: 5.541 min
Delta R.T.: -0.013 min
Response: 914532
Conc: 13.22 ng/ml




#16 AR-1242-1

R.T.: 5.847 min
Delta R.T.: -0.006 min
Response: 619164
Conc: 2.03 ng/ml



#16 AR-1242-1


R.T.: 5.100 min
Delta R.T.: -0.002 min
Response: 139142
Conc: 0.77 ng/ml



#17 AR-1242-2


R.T.: 5.873 min
Delta R.T.: -0.004 min
Response: 289984
Conc: 0.63 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#17 AR-1242-2

R.T.: 5.100 min
Delta R.T.: -0.021 min
Response: 139142
Conc: 0.43 ng/ml




#18 AR-1242-3

R.T.: 5.930 min
Delta R.T.: -0.009 min
Response: 111925
Conc: 0.41 ng/ml



#18 AR-1242-3


R.T.: 5.303 min
Delta R.T.: 0.003 min
Response: 144880
Conc: 1.02 ng/ml



#19 AR-1242-4


R.T.: 6.056 min
Delta R.T.: 0.016 min
Response: 507594
Conc: 2.28 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#19 AR-1242-4

R.T.: 5.375 min
Delta R.T.: -0.006 min
Response: 1508282
Conc: 9.93 ng/ml




#20 AR-1242-5

R.T.: 6.775 min
Delta R.T.: -0.002 min
Response: 128009
Conc: 0.57 ng/ml



#20 AR-1242-5


R.T.: 5.919 min
Delta R.T.: 0.012 min
Response: 1968552
Conc: 10.73 ng/ml



#21 AR-1248-1


R.T.: 5.847 min
Delta R.T.: -0.006 min
Response: 619164
Conc: 2.74 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#21 AR-1248-1

R.T.: 5.100 min
Delta R.T.: -0.003 min
Response: 139142
Conc: 0.91 ng/ml




#22 AR-1248-2

R.T.: 6.110 min
Delta R.T.: -0.012 min
Response: 221802
Conc: 0.67 ng/ml



#22 AR-1248-2


R.T.: 5.342 min
Delta R.T.: 0.005 min
Response: 730387
Conc: 3.21 ng/ml



#23 AR-1248-3


R.T.: 6.335 min
Delta R.T.: 0.004 min
Response: 132697
Conc: 0.37 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#23 AR-1248-3

R.T.: 5.375 min
Delta R.T.: -0.007 min
Response: 1508282
Conc: 6.04 ng/ml




#24 AR-1248-4

R.T.: 6.734 min
Delta R.T.: -0.003 min
Response: 341920
Conc: 0.88 ng/ml



#24 AR-1248-4


R.T.: 5.541 min
Delta R.T.: -0.014 min
Response: 914532
Conc: 3.27 ng/ml



#25 AR-1248-5


R.T.: 6.775 min
Delta R.T.: 0.000 min
Response: 128009
Conc: 0.34 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#25 AR-1248-5

R.T.: 5.956 min
Delta R.T.: 0.006 min
Response: 543200
Conc: 1.93 ng/ml




#26 AR-1254-1

R.T.: 6.705 min
Delta R.T.: -0.003 min
Response: 289217
Conc: 0.71 ng/ml



#26 AR-1254-1


R.T.: 5.919 min
Delta R.T.: 0.011 min
Response: 1968552
Conc: 4.57 ng/ml



#27 AR-1254-2


R.T.: 6.960 min
Delta R.T.: 0.035 min
Response: 307734
Conc: 0.50 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#27 AR-1254-2

R.T.: 6.067 min
Delta R.T.: 0.014 min
Response: 1323130
Conc: 3.49 ng/ml




#28 AR-1254-3

R.T.: 7.286 min
Delta R.T.: -0.009 min
Response: 144669
Conc: 0.23 ng/ml



#28 AR-1254-3


R.T.: 6.463 min
Delta R.T.: 0.002 min
Response: 1851940
Conc: 3.15 ng/ml



#29 AR-1254-4


R.T.: 7.580 min
Delta R.T.: -0.002 min
Response: 346859
Conc: 0.80 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#29 AR-1254-4

R.T.: 6.669 min
Delta R.T.: -0.018 min
Response: 778757
Conc: 1.90 ng/ml




#30 AR-1254-5

R.T.: 7.998 min
Delta R.T.: -0.002 min
Response: 113863
Conc: 0.25 ng/ml



#30 AR-1254-5


R.T.: 7.104 min
Delta R.T.: -0.002 min
Response: 1194854
Conc: 2.52 ng/ml



#31 AR-1260-1


R.T.: 7.459 min
Delta R.T.: 0.003 min
Response: 697368
Conc: 1.82 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#31 AR-1260-1

R.T.: 6.569 min
Delta R.T.: -0.019 min
Response: 1550852
Conc: 4.73 ng/ml




#32 AR-1260-2

R.T.: 7.708 min
Delta R.T.: -0.006 min
Response: 404236
Conc: 0.90 ng/ml



#32 AR-1260-2


R.T.: 6.783 min
Delta R.T.: 0.009 min
Response: 464594
Conc: 1.19 ng/ml



#33 AR-1260-3


R.T.: 8.096 min
Delta R.T.: 0.023 min
Response: 194078
Conc: 0.62 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#33 AR-1260-3

R.T.: 6.924 min
Delta R.T.: -0.007 min
Response: 956241
Conc: 2.71 ng/ml




#34 AR-1260-4

R.T.: 8.318 min
Delta R.T.: 0.007 min
Response: 122494
Conc: 0.35 ng/ml



#34 AR-1260-4


R.T.: 7.387 min
Delta R.T.: -0.015 min
Response: 306757
Conc: 1.11 ng/ml



#35 AR-1260-5


R.T.: 8.652 min
Delta R.T.: 0.002 min
Response: 324406
Conc: 0.47 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#35 AR-1260-5

R.T.: 7.641 min
Delta R.T.: 0.002 min
Response: 773840
Conc: 1.21 ng/ml




#36 AR-1262-1

R.T.: 8.096 min
Delta R.T.: 0.026 min
Response: 194078
Conc: 0.37 ng/ml



#36 AR-1262-1


R.T.: 7.127 min
Delta R.T.: -0.013 min
Response: 1273248
Conc: 2.57 ng/ml



#37 AR-1262-2


R.T.: 8.652 min
Delta R.T.: 0.005 min
Response: 324406
Conc: 0.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#37 AR-1262-2

R.T.: 7.641 min
Delta R.T.: 0.001 min
Response: 773840
Conc: 0.96 ng/ml




#38 AR-1262-3

R.T.: 8.991 min
Delta R.T.: 0.020 min
Response: 148735
Conc: 0.40 ng/ml



#38 AR-1262-3


R.T.: 7.928 min
Delta R.T.: 0.005 min
Response: 539174
Conc: 1.77 ng/ml



#39 AR-1262-4


R.T.: 9.067 min
Delta R.T.: 0.000 min
Response: 62586
Conc: 0.27 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#39 AR-1262-4

R.T.: 7.989 min
Delta R.T.: 0.000 min
Response: 331464
Conc: 0.59 ng/ml




#40 AR-1262-5

R.T.: 9.772 min
Delta R.T.: 0.002 min
Response: 149017
Conc: 0.52 ng/ml



#40 AR-1262-5


R.T.: 8.508 min
Delta R.T.: 0.002 min
Response: 811987
Conc: 3.44 ng/ml



#41 AR-1268-1


R.T.: 8.991 min
Delta R.T.: 0.019 min
Response: 148735
Conc: 0.14 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#41 AR-1268-1

R.T.: 7.928 min
Delta R.T.: 0.005 min
Response: 539174
Conc: 0.59 ng/ml




#42 AR-1268-2

R.T.: 9.075 min
Delta R.T.: 0.002 min
Response: 45980
Conc: 0.05 ng/ml



#42 AR-1268-2


R.T.: 7.989 min
Delta R.T.: 0.000 min
Response: 331464
Conc: 0.40 ng/ml



#43 AR-1268-3


R.T.: 9.337 min
Delta R.T.: 0.027 min
Response: 113087
Conc: 0.14 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK




#43 AR-1268-3

R.T.: 8.193 min
Delta R.T.: -0.006 min
Response: 1041223
Conc: 1.50 ng/ml




#44 AR-1268-4

R.T.: 9.772 min
Delta R.T.: 0.001 min
Response: 149017
Conc: 0.46 ng/ml



#44 AR-1268-4


R.T.: 8.508 min
Delta R.T.: 0.000 min
Response: 811987
Conc: 2.98 ng/ml



#45 AR-1268-5

R.T.: 10.248 min
Delta R.T.: 0.042 min
Response: 104204
Conc: 0.04 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GCP2-BLANK



#45 AR-1268-5

R.T.: 8.850 min
Delta R.T.: 0.028 min
Response: 285168
Conc: 0.14 ng/ml