

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ092120.M
 Title : GC EXTRACTABLES
 Last Update : Tue Sep 22 04:20:45 2020
 Response Via : Initial Calibration

Calibration Files

500 =PQ049893.D 1000 =PQ049891.D 750 =PQ049892.D
 250 =PQ049894.D 50 =PQ049895.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.510	1.419	1.475	1.559	1.781	1.549	E7	9.00
2) SA	Decachlorobiphenyl	6.943	6.038	6.408	7.510	8.350	7.050	E6	12.98
3) L1	AR-1016-1	3.732	3.086	3.295	3.913	4.253	3.656	E5	12.86
4) L1	AR-1016-2	5.716	4.705	4.998	5.922	7.041	5.677	E5	16.07
5) L1	AR-1016-3	3.872	3.147	3.402	3.871	4.617	3.782	E5	14.85
6) L1	AR-1016-4	3.060	2.644	2.850	2.970	3.264	2.958	E5	7.82
7) L1	AR-1016-5	2.647	2.309	2.514	2.638	2.543	2.530	E5	5.40
8) L2	AR-1221-1	1.756					1.756	E5	0.00
9) L2	AR-1221-2	1.309					1.309	E5	0.00
10) L2	AR-1221-3	3.611					3.611	E5	0.00
11) L3	AR-1232-1	3.083					3.083	E5	0.00
12) L3	AR-1232-2	1.550					1.550	E5	0.00
13) L3	AR-1232-3	2.914					2.914	E5	0.00
14) L3	AR-1232-4	1.870					1.870	E5	0.00
15) L3	AR-1232-5	1.404					1.404	E5	0.00
16) L4	AR-1242-1	2.940					2.940	E5	0.00
17) L4	AR-1242-2	4.417					4.417	E5	0.00
18) L4	AR-1242-3	2.974					2.974	E5	0.00
19) L4	AR-1242-4	2.305					2.305	E5	0.00
20) L4	AR-1242-5	2.119					2.119	E5	0.00
21) L5	AR-1248-1	2.318					2.318	E5	0.00
22) L5	AR-1248-2	3.415					3.415	E5	0.00
23) L5	AR-1248-3	3.350					3.350	E5	0.00
24) L5	AR-1248-4	3.614					3.614	E5	0.00
25) L5	AR-1248-5	3.438					3.438	E5	0.00
26) L6	AR-1254-1	3.550					3.550	E5	0.00
27) L6	AR-1254-2	4.680					4.680	E5	0.00
28) L6	AR-1254-3	5.457					5.457	E5	0.00
29) L6	AR-1254-4	4.213					4.213	E5	0.00
30) L6	AR-1254-5	3.719					3.719	E5	0.00
31) L7	AR-1260-1	3.843	3.166	3.454	3.491	4.545	3.700	E5	14.32
32) L7	AR-1260-2	4.283	3.565	3.918	4.208	5.220	4.239	E5	14.55
33) L7	AR-1260-3	3.273	2.730	2.988	3.136	3.984	3.222	E5	14.62
34) L7	AR-1260-4	3.544	2.984	3.250	3.415	3.747	3.388	E5	8.56
35) L7	AR-1260-5	7.043	6.058	6.501	6.686	8.252	6.908	E5	12.03
36) L8	AR-1262-1	4.083					4.083	E5	0.00
37) L8	AR-1262-2	6.913					6.913	E5	0.00
38) L8	AR-1262-3	3.233					3.233	E5	0.00
39) L8	AR-1262-4	3.915					3.915	E5	0.00
40) L8	AR-1262-5	2.953					2.953	E5	0.00
41) L9	AR-1268-1	8.844					8.844	E5	0.00
42) L9	AR-1268-2	8.844					8.844	E5	0.00
43) L9	AR-1268-3	6.903					6.903	E5	0.00
44) L9	AR-1268-4	3.355					3.355	E5	0.00
45) L9	AR-1268-5	2.159					2.159	E6	0.00

Signal #2 Calibration Files

500 =PQ049893.D 1000 =PQ049891.D 750 =PQ049892.D
 250 =PQ049894.D 50 =PQ049895.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	5.881	5.756	5.880	5.828	5.962	5.862	E6	1.30
2) SA	Decachlorobiphenyl	4.782	4.356	4.559	4.910	5.354	4.792	E6	7.90
3) L1	AR-1016-1	2.226	2.063	2.177	2.252	2.523	2.248	E5	7.56
4) L1	AR-1016-2	3.058	2.831	3.000	3.041	3.320	3.050	E5	5.77
5) L1	AR-1016-3	1.533	1.406	1.481	1.516	1.813	1.550	E5	10.02

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ092120.M
 Title : GC EXTRACTABLES
 Last Update : Tue Sep 22 04:20:45 2020
 Response Via : Initial Calibration

Calibration Files

500 =PQ049893.D 1000 =PQ049891.D 750 =PQ049892.D
 250 =PQ049894.D 50 =PQ049895.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	1.208	1.092	1.163	1.208	1.455	1.225	E5	11.19
7)	L1 AR-1016-5	1.604	1.449	1.533	1.648	1.745	1.596	E5	7.04
8)	L2 AR-1221-1	5.910					5.910	E4	0.00
9)	L2 AR-1221-2	4.520					4.520	E4	0.00
10)	L2 AR-1221-3	1.438					1.438	E5	0.00
11)	L3 AR-1232-1	1.196					1.196	E5	0.00
12)	L3 AR-1232-2	1.258					1.258	E5	0.00
13)	L3 AR-1232-3	6.357					6.357	E4	0.00
14)	L3 AR-1232-4	6.006					6.006	E4	0.00
15)	L3 AR-1232-5	6.184					6.184	E4	0.00
16)	L4 AR-1242-1	1.623					1.623	E5	0.00
17)	L4 AR-1242-2	2.200					2.200	E5	0.00
18)	L4 AR-1242-3	1.136					1.136	E5	0.00
19)	L4 AR-1242-4	1.128					1.128	E5	0.00
20)	L4 AR-1242-5	1.558					1.558	E5	0.00
21)	L5 AR-1248-1	1.234					1.234	E5	0.00
22)	L5 AR-1248-2	1.555					1.555	E5	0.00
23)	L5 AR-1248-3	1.654					1.654	E5	0.00
24)	L5 AR-1248-4	1.988					1.988	E5	0.00
25)	L5 AR-1248-5	2.140					2.140	E5	0.00
26)	L6 AR-1254-1	3.157					3.157	E5	0.00
27)	L6 AR-1254-2	2.590					2.590	E5	0.00
28)	L6 AR-1254-3	4.239					4.239	E5	0.00
29)	L6 AR-1254-4	2.966					2.966	E5	0.00
30)	L6 AR-1254-5	3.603					3.603	E5	0.00
31)	L7 AR-1260-1	2.678	2.370	2.546	2.458	2.811	2.573	E5	6.82
32)	L7 AR-1260-2	3.817	3.313	3.616	3.483	4.355	3.717	E5	10.81
33)	L7 AR-1260-3	3.331	2.905	3.137	3.149	4.622	3.429	E5	19.95
34)	L7 AR-1260-4	2.759	2.380	2.571	2.543	3.400	2.731	E5	14.56
35)	L7 AR-1260-5	7.354	6.454	6.916	6.730	7.975	7.086	E5	8.40
36)	L8 AR-1262-1	1.867					1.867	E5	0.00
37)	L8 AR-1262-2	7.204					7.204	E5	0.00
38)	L8 AR-1262-3	2.848					2.848	E5	0.00
39)	L8 AR-1262-4	4.884					4.884	E5	0.00
40)	L8 AR-1262-5	2.423					2.423	E5	0.00
41)	L9 AR-1268-1	9.068					9.068	E5	0.00
42)	L9 AR-1268-2	7.663					7.663	E5	0.00
43)	L9 AR-1268-3	6.792					6.792	E5	0.00
44)	L9 AR-1268-4	2.766					2.766	E5	0.00
45)	L9 AR-1268-5	1.792					1.792	E6	0.00

(#) = Out of Range