

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092222\
 Data File : PQ059247.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2022 22:16
 Operator : YP\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12543176

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 00:03:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091922CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 06:48:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.671	4.006	281.0E6	171.3E6	24.060	22.343
2) SA Decachlor...	10.562	9.097	212.5E6	166.3E6	41.188	39.924
Target Compounds						
26) L6 AR-1254-1	6.707	5.903	175.4E6	168.7E6	431.421	391.940
27) L6 AR-1254-2	6.923	6.048	262.3E6	148.8E6	430.063	393.005
28) L6 AR-1254-3	7.295	6.456	263.9E6	231.3E6	423.743	393.140
29) L6 AR-1254-4	7.582	6.684	177.1E6	160.3E6	410.713	390.802
30) L6 AR-1254-5	8.001	7.101	187.2E6	182.5E6	418.666	385.353

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092222\
Data File : PQ059247.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2022 22:16
Operator : YP\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12543176

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 00:03:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091922CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 06:48:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

