

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092321\
 Data File : PQ054688.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2021 22:58
 Operator : AJ\MA
 Sample : M3872-01
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 SP-LPA-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 06:48:22 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 15 05:25:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.201	4.211	294.1E6	214.6E6	12.314	16.232m#
2) SA Decachlor...	11.364	9.534	289.0E6	194.6E6	17.125	14.356
Target Compounds						
16) L4 AR-1242-1	6.523	5.466	172.7E6	94355391	293.187	292.495
17) L4 AR-1242-2	6.547	5.486	232.3E6	156.4E6	258.944	306.889
18) L4 AR-1242-3	6.614	5.679	107.8E6	59786903	191.787	232.834m
19) L4 AR-1242-4	6.722	5.775	103.9E6	65574341	227.133	259.096
20) L4 AR-1242-5	7.502	6.338	78102383	64812960	166.804	182.429m



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092321\
Data File : PQ054688.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2021 22:58
Operator : AJ\MA
Sample : M3872-01
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SP-LPA-01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 06:48:22 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091421.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 15 05:25:53 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.201 min
Delta R.T.: 0.000 min
Response: 294077264
Conc: 12.31 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SP-LPA-01




#1 Tetrachloro-m-xylene

R.T.: 4.211 min
Delta R.T.: 0.003 min
Response: 214588245
Conc: 16.23 ng/ml m




#2 Decachlorobiphenyl

R.T.: 11.364 min
Delta R.T.: -0.013 min
Response: 289045383
Conc: 17.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.534 min
Delta R.T.: -0.004 min
Response: 194553687
Conc: 14.36 ng/ml



#16 AR-1242-1


R.T.: 6.523 min
Delta R.T.: -0.003 min
Response: 172701052
Conc: 293.19 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SP-LPA-01




#16 AR-1242-1

R.T.: 5.466 min
Delta R.T.: -0.001 min
Response: 94355391
Conc: 292.49 ng/ml




#17 AR-1242-2

R.T.: 6.547 min
Delta R.T.: -0.003 min
Response: 232333499
Conc: 258.94 ng/ml



#17 AR-1242-2


R.T.: 5.486 min
Delta R.T.: -0.001 min
Response: 156397557
Conc: 306.89 ng/ml



#18 AR-1242-3


R.T.: 6.614 min
Delta R.T.: -0.002 min
Response: 107774753
Conc: 191.79 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SP-LPA-01




#18 AR-1242-3

R.T.: 5.679 min
Delta R.T.: -0.001 min
Response: 59786903
Conc: 232.83 ng/ml m




#19 AR-1242-4

R.T.: 6.722 min
Delta R.T.: -0.003 min
Response: 103909220
Conc: 227.13 ng/ml



#19 AR-1242-4


R.T.: 5.775 min
Delta R.T.: 0.000 min
Response: 65574341
Conc: 259.10 ng/ml



#20 AR-1242-5

R.T.: 7.502 min
Delta R.T.: -0.003 min
Response: 78102383
Conc: 166.80 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SP-LPA-01



#20 AR-1242-5

R.T.: 6.338 min
Delta R.T.: -0.001 min
Response: 64812960
Conc: 182.43 ng/ml m