

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ092419CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Sep 24 07:41:41 2019
 Response Via : Initial Calibration

Calibration Files

400 =PQ043773.D 1600 =PQ043775.D 800 =PQ043774.D
 200 =PQ043772.D 100 =PQ043771.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.591	1.904	1.992	1.960	1.932	1.876	E6	8.67
2) SA	Decachlorobiphenyl	6.144	5.747	6.192	6.640	6.876	6.320	E6	7.02
3) L1	AR-1016-1	7.302	7.846	8.256	8.676	8.209	8.058	E4	6.39
4) L1	AR-1016-2	1.030	1.130	1.199	1.188	1.227	1.155	E5	6.76
5) L1	AR-1016-3	6.226	6.781	7.202	7.380	7.220	6.962	E4	6.71
6) L1	AR-1016-4	5.336	5.764	6.110	6.462	6.406	6.016	E4	7.83
7) L1	AR-1016-5	5.825	5.824	6.286	6.681	7.193	6.362	E4	9.22
8) L2	AR-1221-1	2.334	2.243	2.360	2.456	2.387	2.356	E4	3.31
9) L2	AR-1221-2	1.777	1.605	1.687	2.052	1.978	1.820	E4	10.44
10) L2	AR-1221-3	5.795	5.625	5.763	6.129	5.631	5.789	E4	3.54
11) L3	AR-1232-1	5.193	4.482	4.927	5.241	5.325	5.034	E4	6.80
12) L3	AR-1232-2	2.749	2.401	2.642	2.914	2.462	2.634	E4	7.96
13) L3	AR-1232-3	5.811	5.207	5.615	5.969	5.544	5.629	E4	5.13
14) L3	AR-1232-4	2.925	2.664	2.818	2.873	2.777	2.811	E4	3.53
15) L3	AR-1232-5	2.208	1.974	2.121	2.390	2.149	2.168	E4	6.97
16) L4	AR-1242-1	7.504	7.279	7.653	8.131	7.888	7.691	E4	4.30
17) L4	AR-1242-2	1.099	1.075	1.122	1.181	1.132	1.122	E5	3.56
18) L4	AR-1242-3	6.671	6.460	6.798	7.272	6.823	6.805	E4	4.38
19) L4	AR-1242-4	5.595	5.465	5.713	6.114	5.985	5.774	E4	4.67
20) L4	AR-1242-5	7.439	7.086	7.520	8.058	8.169	7.654	E4	5.90
21) L5	AR-1248-1	6.839	6.171	6.308	7.038	6.577	6.586	E4	5.46
22) L5	AR-1248-2	0.990	0.874	0.901	1.048	0.952	0.953	E5	7.29
23) L5	AR-1248-3	1.140	1.019	1.043	1.172	1.098	1.094	E5	5.86
24) L5	AR-1248-4	1.439	1.312	1.345	1.501	1.389	1.397	E5	5.38
25) L5	AR-1248-5	1.372	1.264	1.300	1.422	1.333	1.338	E5	4.59
26) L6	AR-1254-1	1.303	1.195	1.328	1.332	1.598	1.351	E5	11.02
27) L6	AR-1254-2	2.111	1.963	2.155	2.163	2.404	2.159	E5	7.36
28) L6	AR-1254-3	2.463	2.339	2.536	2.524	2.847	2.542	E5	7.39
29) L6	AR-1254-4	1.991	1.885	2.045	2.001	2.166	2.017	E5	5.04
30) L6	AR-1254-5	2.604	2.494	2.695	2.633	2.830	2.651	E5	4.67
31) L7	AR-1260-1	1.269	1.280	1.372	1.418	1.441	1.356	E5	5.78
32) L7	AR-1260-2	1.657	1.690	1.821	1.932	1.833	1.787	E5	6.31
33) L7	AR-1260-3	1.397	1.404	1.512	1.568	1.564	1.489	E5	5.62
34) L7	AR-1260-4	1.728	1.741	1.869	1.948	1.934	1.844	E5	5.67
35) L7	AR-1260-5	3.965	4.068	4.305	4.362	4.399	4.220	E5	4.55
36) L8	AR-1262-1	2.762	2.682	2.786	3.025	2.861	2.823	E5	4.58
37) L8	AR-1262-2	5.922	5.951	6.060	6.308	6.046	6.058	E5	2.51
38) L8	AR-1262-3	4.286	4.143	4.277	4.599	4.406	4.342	E5	3.94
39) L8	AR-1262-4	3.264	3.158	3.264	3.506	3.426	3.323	E5	4.20
40) L8	AR-1262-5	2.776	2.667	2.751	2.979	2.915	2.818	E5	4.51
41) L9	AR-1268-1	7.808	7.313	7.534	7.534	8.212	7.680	E5	4.49
42) L9	AR-1268-2	7.437	7.027	7.190	7.168	7.688	7.302	E5	3.58
43) L9	AR-1268-3	6.309	5.923	6.079	6.085	6.425	6.164	E5	3.26
44) L9	AR-1268-4	3.264	2.988	3.116	3.172	3.501	3.208	E5	5.98
45) L9	AR-1268-5	2.596	2.472	2.539	2.492	2.634	2.547	E6	2.68

Signal #2 Calibration Files

400 =PQ043773.D 1600 =PQ043775.D 800 =PQ043774.D
 200 =PQ043772.D 100 =PQ043771.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	0.851	0.992	1.043	1.018	1.073	0.995	E6	8.65
2) SA	Decachlorobiphenyl	2.048	1.928	2.067	2.232	2.254	2.106	E6	6.48
3) L1	AR-1016-1	3.361	3.545	3.785	3.905	3.824	3.684	E4	6.10
4) L1	AR-1016-2	4.590	4.912	5.578	5.416	5.401	5.180	E4	7.98
5) L1	AR-1016-3	2.399	2.553	2.718	2.809	2.830	2.662	E4	6.88

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ092419CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Sep 24 07:41:41 2019
 Response Via : Initial Calibration

Calibration Files

400 =PQ043773.D 1600 =PQ043775.D 800 =PQ043774.D
 200 =PQ043772.D 100 =PQ043771.D

	Compound	400	1600	800	200	100	Avg		%RSD
6)	L1 AR-1016-4	1.985	2.062	2.222	2.337	2.280	2.177	E4	6.83
7)	L1 AR-1016-5	2.593	2.677	2.870	2.925	3.051	2.823	E4	6.60
8)	L2 AR-1221-1	1.348	1.219	1.288	1.607	1.345	1.361	E4	10.78
9)	L2 AR-1221-2	0.902	0.873	0.941	1.152	0.978	0.969	E4	11.31
10)	L2 AR-1221-3	3.105	2.970	3.086	3.287	3.415	3.173	E4	5.57
11)	L3 AR-1232-1	2.756	2.370	2.598	2.864	3.184	2.755	E4	11.02
12)	L3 AR-1232-2	2.557	2.291	2.488	2.569	2.401	2.461	E4	4.72
13)	L3 AR-1232-3	1.304	1.178	1.261	1.341	1.307	1.278	E4	4.91
14)	L3 AR-1232-4	1.203	1.049	1.149	1.255	1.248	1.181	E4	7.18
15)	L3 AR-1232-5	1.298	1.160	1.225	1.272	1.402	1.271	E4	7.06
16)	L4 AR-1242-1	3.406	3.360	3.514	3.746	3.441	3.493	E4	4.35
17)	L4 AR-1242-2	4.815	4.600	4.875	5.181	4.911	4.877	E4	4.28
18)	L4 AR-1242-3	2.467	2.418	2.528	2.701	2.643	2.551	E4	4.65
19)	L4 AR-1242-4	2.495	2.390	2.518	2.801	2.672	2.575	E4	6.28
20)	L4 AR-1242-5	3.465	3.245	3.456	3.835	3.355	3.471	E4	6.40
21)	L5 AR-1248-1	3.125	2.850	2.901	3.174	2.934	2.997	E4	4.79
22)	L5 AR-1248-2	4.155	3.655	3.788	4.023	4.061	3.936	E4	5.27
23)	L5 AR-1248-3	4.283	3.784	3.905	4.378	4.270	4.124	E4	6.35
24)	L5 AR-1248-4	5.143	4.620	4.741	5.531	4.883	4.983	E4	7.28
25)	L5 AR-1248-5	5.409	4.842	4.998	5.868	5.437	5.311	E4	7.61
26)	L6 AR-1254-1	7.603	6.949	7.712	7.898	7.648	7.562	E4	4.77
27)	L6 AR-1254-2	6.797	6.106	6.813	7.252	7.992	6.992	E4	9.92
28)	L6 AR-1254-3	1.162	1.082	1.181	1.195	1.250	1.174	E5	5.22
29)	L6 AR-1254-4	7.968	7.362	8.067	8.207	8.309	7.983	E4	4.64
30)	L6 AR-1254-5	1.074	1.007	1.099	1.093	1.167	1.088	E5	5.27
31)	L7 AR-1260-1	5.408	5.340	5.828	6.199	6.407	5.836	E4	8.07
32)	L7 AR-1260-2	6.955	6.870	7.589	8.050	8.256	7.544	E4	8.29
33)	L7 AR-1260-3	6.430	6.413	6.860	7.358	7.380	6.888	E4	6.89
34)	L7 AR-1260-4	5.597	5.526	6.007	6.314	6.302	5.949	E4	6.31
35)	L7 AR-1260-5	1.441	1.464	1.545	1.616	1.593	1.532	E5	5.03
36)	L8 AR-1262-1	4.497	4.212	4.421	4.889	4.551	4.514	E4	5.45
37)	L8 AR-1262-2	2.055	2.053	2.083	2.204	2.084	2.096	E5	2.96
38)	L8 AR-1262-3	8.887	8.446	8.468	9.776	8.841	8.884	E4	6.07
39)	L8 AR-1262-4	1.669	1.636	1.687	1.794	1.718	1.701	E5	3.52
40)	L8 AR-1262-5	1.027	0.961	0.998	1.075	1.047	1.021	E5	4.32
41)	L9 AR-1268-1	2.651	2.457	2.540	2.550	2.751	2.590	E5	4.39
42)	L9 AR-1268-2	2.497	2.334	2.403	2.396	2.582	2.443	E5	3.99
43)	L9 AR-1268-3	2.177	2.009	2.075	2.116	2.262	2.128	E5	4.54
44)	L9 AR-1268-4	1.190	1.070	1.124	1.163	1.297	1.169	E5	7.23
45)	L9 AR-1268-5	9.094	8.687	8.890	8.632	9.135	8.888	E5	2.58

(#) = Out of Range