Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092419\
Data File : PQ043786.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2019 00:24
Operator : SM\AJ
Sample : AR1242ICC100
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242101

Manual Integrations APPROVED
Ankita
9/24/2019 3:24:00 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 03:41:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 24 03:41:12 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	5.230	4.406	11175316	5832652	4.946m	4.917m
2)	SA Decachlor...	11.398	9.861	74080651	24946495	10.748	10.721
	Target Compounds						
16)	L4 AR-1242-1	6.545	5.689	7887540	3440879	105.111	101.024
17)	L4 AR-1242-2	6.569	5.711	11320687	4910711	103.044	101.987
18)	L4 AR-1242-3	6.636	5.908	6823094	2643099	102.283	107.142
19)	L4 AR-1242-4	6.743	6.001	5985043	2672331	106.970	107.118
20)	L4 AR-1242-5	7.526	6.572	8169414	3355429	109.814	96.830

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092419\
Data File : PQ043786.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2019 00:24
Operator : SM\AJ
Sample : AR1242ICC100
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242101

Manual Integrations APPROVED
Ankita
9/24/2019 3:24:00 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 03:41:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 24 03:41:12 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm