

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092421\
 Data File : PQ054756.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Sep 2021 21:04
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 144 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 01:08:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 15 05:25:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.200	4.207	1321.2E6	722.4E6	55.322	54.643
2)	SA Decachlor...	11.363	9.532	851.3E6	687.0E6	50.436	50.697
Target Compounds							
26)	L6 AR-1254-1	7.431	6.336	395.8E6	374.4E6	454.383	491.479
27)	L6 AR-1254-2	7.658	6.495	612.9E6	344.0E6	458.202	491.541
28)	L6 AR-1254-3	8.042	6.915	637.0E6	511.3E6	457.348	463.800
29)	L6 AR-1254-4	8.338	7.155	489.2E6	358.3E6	459.209	473.808
30)	L6 AR-1254-5	8.765	7.582	522.8E6	531.2E6	454.873	479.942

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092421\
Data File : PQ054756.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2021 21:04
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 144 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 01:08:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091421.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 15 05:25:53 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

