

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092520\
 Data File : PQ050020.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Sep 2020 11:29
 Operator : DD\AJ
 Sample : L1955-08
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 MDL-WATER-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 05:48:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092420CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 24 08:00:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.214	4.271	170.2E6	59853040	17.194	16.979
2)	SA Decachlor...	11.410	9.621	275.0E6	240.9E6	40.512	44.790
Target Compounds							
3)	L1 AR-1016-1	6.544	5.531	10484511	4205512	32.448	25.302
4)	L1 AR-1016-2	6.568	5.551	14562408	5291084	30.978	23.684
5)	L1 AR-1016-3	6.635	5.744	9139426	3131287	30.106	26.428
6)	L1 AR-1016-4	6.744	5.795	7724979	2726902	31.773	28.894
7)	L1 AR-1016-5	7.058	6.027	7133562	2467098	32.312	19.956 #
31)	L7 AR-1260-1	8.235	7.121	11721334	7507342	35.991	30.636
32)	L7 AR-1260-2	8.501	7.316	12532147	10352091	32.380	29.353
33)	L7 AR-1260-3	8.867	7.473	9720776	9722337	32.374	32.265
34)	L7 AR-1260-4	9.111	7.955	9827123	7472093	30.483	27.405
35)	L7 AR-1260-5	9.459	8.202	21069590	19886170	31.398	25.798

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092520\
Data File : PQ050020.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2020 11:29
Operator : DD\AJ
Sample : L1955-08
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleID :
MDL-WATER-01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 05:48:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092420CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 24 08:00:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

