

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092523\
 Data File : PQ063327.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Sep 2023 17:28
 Operator : YP\AJ
 Sample : AR1232ICC1600
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12325013

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 05:30:12 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092523CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 26 05:27:42 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.462	2.781	392.1E6	297.2E6	73.730	74.766
2) SA Decachlor...	8.697	7.582	539.9E6	519.0E6	140.896	141.370
Target Compounds						
11) L3 AR-1232-1	3.815	3.131	148.8E6	122.7E6	1345.494	1310.556
12) L3 AR-1232-2	4.316	3.808	82743332	121.7E6	1388.782	1413.775
13) L3 AR-1232-3	4.597	3.969	162.3E6	63957244	1422.304	1414.595
14) L3 AR-1232-4	4.747	4.055	79994858	54350059	1387.585	1396.223
15) L3 AR-1232-5	4.846	4.214	56829463	63660625	1387.864	1382.080

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092523\
Data File : PQ063327.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Sep 2023 17:28
Operator : YP\AJ
Sample : AR1232ICC1600
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12325013

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 05:30:12 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092523CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 26 05:27:42 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

