

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092524\
 Data File : PQ068858.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Sep 2024 07:39
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/26/2024
 Supervised By :Ankita Jodhani 09/26/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 07:59:26 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 13:06:53 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.437	2.818	337.4E6	391.8E6	47.255	55.011
2) SA Decachlor...	8.609	7.592	180.6E6	314.9E6	41.430	47.677
Target Compounds						
3) L1 AR-1016-1	4.537	3.822	108.4E6	133.5E6	463.907	562.050
4) L1 AR-1016-2	4.556	3.837	166.2E6	200.4E6	452.722	554.677
5) L1 AR-1016-3	4.613	3.998	104.4E6	108.1E6	462.053	556.685
6) L1 AR-1016-4	4.705	4.047	86763522	91003346	455.706	563.498
7) L1 AR-1016-5	4.993	4.242	81357339	108.9E6	460.149	563.952m
31) L7 AR-1260-1	6.092	5.239	140.5E6	206.4E6	422.629	537.599 #
32) L7 AR-1260-2	6.351	5.429	157.8E6	247.1E6	421.026	542.104 #
33) L7 AR-1260-3	6.705	5.571	98126190	232.2E6	417.539	532.763 #
34) L7 AR-1260-4	6.923	6.032	115.8E6	167.3E6	417.885	520.749
35) L7 AR-1260-5	7.233	6.278	213.3E6	367.7E6	399.147	504.630 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092524\
Data File : PQ068858.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Sep 2024 07:39
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/26/2024
Supervised By :Ankita Jodhani 09/26/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 07:59:26 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 13:06:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

