

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092622\
 Data File : PQ059323.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2022 07:45
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR16603180

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/27/2022
 Supervised By :Ankita Jodhani 09/27/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 08:04:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092622CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 27 05:37:36 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.322	2.727	322.9E6	144.9E6	18.267	16.937m
2)	SA Decachlor...	8.419	7.438	296.6E6	236.3E6	38.602	36.663
Target Compounds							
3)	L1 AR-1016-1	4.411	3.713	163.9E6	61099547	352.269	317.360m
4)	L1 AR-1016-2	4.430	3.727	297.0E6	157.2E6	356.326	336.873m
5)	L1 AR-1016-3	4.485	3.884	202.8E6	68513806	380.104	336.546
6)	L1 AR-1016-4	4.578	3.928	154.8E6	66943589	369.760	364.790
7)	L1 AR-1016-5	4.855	4.120	143.6E6	74572945	367.607	343.367
31)	L7 AR-1260-1	5.941	5.096	274.2E6	165.2E6	373.324	360.384
32)	L7 AR-1260-2	6.200	5.285	314.3E6	197.4E6	368.762	359.086
33)	L7 AR-1260-3	6.549	5.424	236.2E6	184.2E6	379.368	361.824
34)	L7 AR-1260-4	6.767	5.882	266.5E6	149.6E6	378.303	358.137
35)	L7 AR-1260-5	7.078	6.129	520.7E6	331.1E6	375.725	352.829

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092622\
Data File : PQ059323.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 07:45
Operator : YP\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR16603180

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/27/2022
Supervised By :Ankita Jodhani 09/27/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 08:04:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092622CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 27 05:37:36 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

