

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092623\
 Data File : PQ063372.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Sep 2023 17:24
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 01:28:52 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 20 05:26:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.463	2.781	241.9E6	178.0E6	51.657	51.510
2) SA Decachlor...	8.697	7.582	178.3E6	170.4E6	51.033	47.410
Target Compounds						
26) L6 AR-1254-1	5.408	4.553	91511704	99699356	455.036	472.842
27) L6 AR-1254-2	5.627	4.701	140.3E6	87049292	459.990	470.328
28) L6 AR-1254-3	5.983	5.085	149.6E6	137.3E6	455.993	479.779
29) L6 AR-1254-4	6.268	5.312	109.1E6	87691413	450.451	473.668
30) L6 AR-1254-5	6.685	5.719	117.9E6	123.6E6	459.040	478.270



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092623\
Data File : PQ063372.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Sep 2023 17:24
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 01:28:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091923.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 20 05:26:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.463 min
Delta R.T.: 0.000 min
Response: 241937008
Conc: 51.66 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.781 min
Delta R.T.: 0.000 min
Response: 178017576
Conc: 51.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.697 min
Delta R.T.: -0.001 min
Response: 178337342
Conc: 51.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.582 min
Delta R.T.: -0.003 min
Response: 170425048
Conc: 47.41 ng/ml



#26 AR-1254-1


R.T.: 5.408 min
Delta R.T.: 0.000 min
Response: 91511704
Conc: 455.04 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 4.553 min
Delta R.T.: 0.000 min
Response: 99699356
Conc: 472.84 ng/ml




#27 AR-1254-2

R.T.: 5.627 min
Delta R.T.: 0.000 min
Response: 140251389
Conc: 459.99 ng/ml



#27 AR-1254-2


R.T.: 4.701 min
Delta R.T.: 0.000 min
Response: 87049292
Conc: 470.33 ng/ml



#28 AR-1254-3


R.T.: 5.983 min
Delta R.T.: 0.000 min
Response: 149584873
Conc: 455.99 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.085 min
Delta R.T.: 0.000 min
Response: 137277939
Conc: 479.78 ng/ml




#29 AR-1254-4

R.T.: 6.268 min
Delta R.T.: 0.000 min
Response: 109102636
Conc: 450.45 ng/ml



#29 AR-1254-4


R.T.: 5.312 min
Delta R.T.: 0.000 min
Response: 87691413
Conc: 473.67 ng/ml



#30 AR-1254-5

R.T.: 6.685 min
Delta R.T.: 0.000 min
Response: 117887004
Conc: 459.04 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 5.719 min
Delta R.T.: -0.001 min
Response: 123569106
Conc: 478.27 ng/ml