

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092723\
 Data File : PQ063415.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2023 16:14
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 01:22:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 20 05:26:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.462	2.781	244.9E6	180.9E6	52.295	52.356
2)	SA Decachlor...	8.698	7.582	179.1E6	168.1E6	51.257	46.761
Target Compounds							
21)	L5 AR-1248-1	4.578	3.792	47010318	37858471	478.376	471.781
22)	L5 AR-1248-2	4.847	4.018	69249106	56266586	484.189	469.946
23)	L5 AR-1248-3	5.038	4.055	85285380	53988055	495.763	470.898
24)	L5 AR-1248-4	5.436	4.215	88840626	66797790	464.671	473.733
25)	L5 AR-1248-5	5.471	4.591	82908180	65946429	445.652	477.245



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092723\
Data File : PQ063415.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2023 16:14
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 01:22:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091923.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 20 05:26:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.462 min
Delta R.T.: 0.000 min
Response: 244926314
Conc: 52.30 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.781 min
Delta R.T.: -0.001 min
Response: 180941415
Conc: 52.36 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.698 min
Delta R.T.: 0.000 min
Response: 179119629
Conc: 51.26 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.582 min
Delta R.T.: -0.003 min
Response: 168094087
Conc: 46.76 ng/ml



#21 AR-1248-1


R.T.: 4.578 min
Delta R.T.: 0.000 min
Response: 47010318
Conc: 478.38 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 3.792 min
Delta R.T.: -0.002 min
Response: 37858471
Conc: 471.78 ng/ml




#22 AR-1248-2

R.T.: 4.847 min
Delta R.T.: 0.000 min
Response: 69249106
Conc: 484.19 ng/ml



#22 AR-1248-2


R.T.: 4.018 min
Delta R.T.: -0.002 min
Response: 56266586
Conc: 469.95 ng/ml



#23 AR-1248-3


R.T.: 5.038 min
Delta R.T.: -0.002 min
Response: 85285380
Conc: 495.76 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.055 min
Delta R.T.: 0.000 min
Response: 53988055
Conc: 470.90 ng/ml




#24 AR-1248-4

R.T.: 5.436 min
Delta R.T.: 0.000 min
Response: 88840626
Conc: 464.67 ng/ml



#24 AR-1248-4


R.T.: 4.215 min
Delta R.T.: 0.000 min
Response: 66797790
Conc: 473.73 ng/ml



#25 AR-1248-5

R.T.: 5.471 min
Delta R.T.: 0.000 min
Response: 82908180
Conc: 445.65 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 4.591 min
Delta R.T.: 0.000 min
Response: 65946429
Conc: 477.25 ng/ml