

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092724\
 Data File : PQ068939.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2024 15:35
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/30/2024
 Supervised By :Ankita Jodhani 09/30/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 04:37:23 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 13:06:53 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.436	2.817	340.9E6	392.5E6	47.735	55.112
2)	SA Decachlor...	8.608	7.590	182.0E6	309.1E6	41.760	46.809
Target Compounds							
3)	L1 AR-1016-1	4.535	3.820	109.7E6	136.4E6	469.426	574.224
4)	L1 AR-1016-2	4.555	3.836	165.9E6	205.0E6	451.810	567.243 #
5)	L1 AR-1016-3	4.612	3.996	104.6E6	110.3E6	462.606	567.806
6)	L1 AR-1016-4	4.704	4.045	86609876	92060522	454.899	570.044 #
7)	L1 AR-1016-5	4.991	4.240	83045194	110.6E6	469.695	572.462m
31)	L7 AR-1260-1	6.091	5.237	147.8E6	204.1E6	444.370	531.571
32)	L7 AR-1260-2	6.351	5.427	165.8E6	242.9E6	442.429	533.021
33)	L7 AR-1260-3	6.704	5.568	102.6E6	225.9E6	436.656	518.400
34)	L7 AR-1260-4	6.922	6.030	122.1E6	165.2E6	440.530m	514.090
35)	L7 AR-1260-5	7.232	6.276	220.9E6	362.1E6	413.373	496.935

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ092724\
Data File : PQ068939.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2024 15:35
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2024
Supervised By :Ankita Jodhani 09/30/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 04:37:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 13:06:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

