

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100218\
 Data File : PQ032515.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2018 18:39
 Operator : SM\SJ
 Sample : J5133-19
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 092618-JETT-F-U-M

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 03 01:53:35 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 19 02:07:33 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.589	3.867	57217034	43949116	24.676	21.945
2) SA Decachlor...	10.444	8.939	35031201	30598339	16.804	15.802m

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100218\
Data File : PQ032515.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Oct 2018 18:39
Operator : SM\SJ
Sample : J5133-19
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
092618-JETT-F-U-M

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 03 01:53:35 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091818.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 19 02:07:33 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.589 min
Delta R.T.: -0.005 min
Response: 57217034
Conc: 24.68 ng/ml

Instrument :
ECD_Q
ClientSampleId :
092618-JETT-F-U-M




#1 Tetrachloro-m-xylene

R.T.: 3.867 min
Delta R.T.: -0.007 min
Response: 43949116
Conc: 21.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.444 min
Delta R.T.: -0.007 min
Response: 35031201
Conc: 16.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.939 min
Delta R.T.: -0.013 min
Response: 30598339
Conc: 15.80 ng/ml m