

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100218\
 Data File : PQ032522.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2018 20:20
 Operator : SM\SJ
 Sample : J5187-06
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 092818-JETT-F-U-S

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 03 02:01:00 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 19 02:07:33 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.590	3.868	40215010	31777518	17.344	15.867
2) SA Decachlor...	10.444	8.939	27332066	24184653	13.111m	12.490m

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100218\
Data File : PQ032522.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Oct 2018 20:20
Operator : SM\SJ
Sample : J5187-06
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
092818-JETT-F-U-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 03 02:01:00 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091818.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 19 02:07:33 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.590 min
Delta R.T.: -0.004 min
Response: 40215010
Conc: 17.34 ng/ml

Instrument :
ECD_Q
ClientSampleId :
092818-JETT-F-U-S




#1 Tetrachloro-m-xylene

R.T.: 3.868 min
Delta R.T.: -0.006 min
Response: 31777518
Conc: 15.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.444 min
Delta R.T.: -0.008 min
Response: 27332066
Conc: 13.11 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.939 min
Delta R.T.: -0.013 min
Response: 24184653
Conc: 12.49 ng/ml m