

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100821\
 Data File : PQ054911.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Oct 2021 13:11
 Operator : AJ\MA
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12483126

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 04:41:21 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100421CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 05 07:08:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.196	4.202	444.8E6	208.2E6	19.208	19.004
2) SA Decachlor...	11.358	9.524	673.0E6	495.6E6	37.630	35.019
Target Compounds						
21) L5 AR-1248-1	6.520	5.460	160.2E6	85800456	374.680	372.671
22) L5 AR-1248-2	6.811	5.721	232.7E6	131.9E6	359.185	366.792
23) L5 AR-1248-3	7.030	5.766	258.6E6	135.6E6	350.756	365.422
24) L5 AR-1248-4	7.459	5.951	298.1E6	164.3E6	345.491	364.998
25) L5 AR-1248-5	7.500	6.374	305.4E6	174.7E6	346.457	358.697

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100821\
Data File : PQ054911.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Oct 2021 13:11
Operator : AJ\MA
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12483126

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 04:41:21 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100421CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 05 07:08:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

