

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100919\
 Data File : PQ044199.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2019 01:04
 Operator : HP\AJ
 Sample : AIBLK44
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AIBLK44

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 04:24:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 10 02:55:10 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.317	4.356	108.9E6	129.0E6	28.014	28.889
2) SA Decachlor...	11.598	9.799	237.1E6	267.9E6	44.250	44.785
Target Compounds						
30) L6 AR-1254-5	8.860	0.000	119.1E6	0	497.755	N.D. #
31) L7 AR-1260-1	8.353	0.000	3072609	0	20.411	N.D. #
32) L7 AR-1260-2	8.604	0.000	9550854	0	48.528	N.D. #
33) L7 AR-1260-3	9.047	0.000	3199037	0	19.953	N.D. #
36) L8 AR-1262-1	9.047	0.000	3199037	0	29.455	N.D. #
43) L9 AR-1268-3	10.279	0.000	1796003	0	3.129	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100919\
Data File : PQ044199.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2019 01:04
Operator : HP\AJ
Sample : AIBLK44
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AIBLK44

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 04:24:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 10 02:55:10 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.317 min
Delta R.T.: 0.000 min
Response: 108909678
Conc: 28.01 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AIBLK44




#1 Tetrachloro-m-xylene

R.T.: 4.356 min
Delta R.T.: 0.001 min
Response: 128956206
Conc: 28.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.598 min
Delta R.T.: 0.000 min
Response: 237072715
Conc: 44.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.799 min
Delta R.T.: -0.009 min
Response: 267877023
Conc: 44.79 ng/ml



#30 AR-1254-5


R.T.: 8.860 min
Delta R.T.: -0.041 min
Response: 119135033
Conc: 497.75 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AIBLK44




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.785 min
Response: 0
Conc: N.D.




#31 AR-1260-1

R.T.: 8.353 min
Delta R.T.: 0.012 min
Response: 3072609
Conc: 20.41 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T. : 7.249 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 8.604 min
Delta R.T.: -0.003 min
Response: 9550854
Conc: 48.53 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AIBLK44




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.444 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 9.047 min
Delta R.T.: 0.067 min
Response: 3199037
Conc: 19.95 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 7.606 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 9.047 min
Delta R.T.: 0.009 min
Response: 3199037
Conc: 29.46 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AIBLK44




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 7.782 min
Response: 0
Conc: N.D.



#43 AR-1268-3

R.T.: 10.279 min
Delta R.T.: 0.000 min
Response: 1796003
Conc: 3.13 ng/ml



#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 8.907 min
Response: 0
Conc: N.D.