

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100920\
 Data File : PQ050332.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2020 23:58
 Operator : DD\AJ
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 142 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248343

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 08:44:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092420CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 24 08:00:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.205	4.267	260.3E6	78218126	26.293	22.189
2) SA Decachlor...	11.390	9.614	278.6E6	201.3E6	41.045	37.426
Target Compounds						
21) L5 AR-1248-1	6.537	5.527	109.7E6	39650455	504.726	387.242
22) L5 AR-1248-2	6.828	5.791	142.7E6	50374914	515.343	386.428 #
23) L5 AR-1248-3	7.049	5.836	151.7E6	50593099	504.783	381.391
24) L5 AR-1248-4	7.480	6.022	174.9E6	62151668	522.918	376.417 #
25) L5 AR-1248-5	7.519	6.444	165.5E6	67931678	525.627	374.795 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100920\
Data File : PQ050332.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2020 23:58
Operator : DD\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 142 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
AR1248343

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 08:44:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092420CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 24 08:00:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

