

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100923\
 Data File : PQ063625.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2023 09:04
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/10/2023
 Supervised By :Ankita Jodhani 10/11/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 10:57:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 20 05:26:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.461	2.779	230.0E6	175.1E6	49.110	50.668
2)	SA Decachlor...	8.696	7.579	161.3E6	190.8E6	46.169	53.086
Target Compounds							
3)	L1 AR-1016-1	4.575	3.790	68809206	59849940	473.255	480.481
4)	L1 AR-1016-2	4.595	3.806	108.8E6	85986356	465.852	483.724
5)	L1 AR-1016-3	4.653	3.966	67835002	47569943	472.072	492.994m
6)	L1 AR-1016-4	4.746	4.016	55593765	40243648	487.376	496.452m
7)	L1 AR-1016-5	5.035	4.212	55777626	50104627	474.670m	497.739m
31)	L7 AR-1260-1	6.147	5.212	115.3E6	95354424	479.344m	472.231m
32)	L7 AR-1260-2	6.407	5.403	123.7E6	113.9E6	436.451m	465.560m
33)	L7 AR-1260-3	6.764	5.545	97312656	108.2E6	478.715	457.729m
34)	L7 AR-1260-4	6.982	6.009	103.1E6	90801430	432.654	475.591
35)	L7 AR-1260-5	7.295	6.256	196.5E6	205.6E6	438.916	469.518

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ100923\
Data File : PQ063625.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2023 09:04
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/10/2023
Supervised By :Ankita Jodhani 10/11/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:57:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091923.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 20 05:26:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

