

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101019\
Data File : PQ044207.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2019 09:44
Operator : HP\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AR1248310

Manual Integrations
APPROVED

Ankita
10/11/2019 1:45:05 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 02:35:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 10 02:55:10 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(21) AR-1248-1 (L5)
R.T. Response Conc
6.64 34637964 472.06
6.94 46049669 464.87
7.16 49442295 457.81
7.59 56982836 452.60
7.63 57536778 475.20

(21) AR-1248-1 #2 (L5)
R.T. Response Conc
5.64 38456031 424.68
5.91 49300742 355.69
5.95 58836310 416.02
6.14 81297839 473.31
6.57 85030313 507.89

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101019\
Data File : PQ044207.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2019 09:44
Operator : HP\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AR1248310

Manual Integrations
APPROVED

Ankita
10/11/2019 1:45:05 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 02:35:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 10 02:55:10 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(21) AR-1248-1 (L5)
R.T. Response Conc
6.64 34637964 472.06
6.94 46049669 464.87
7.16 49442295 457.81
7.59 56982836 452.60
7.63 57536778 475.20

(21) AR-1248-1 #2 (L5)
R.T. Response Conc
5.64 38456031 424.68
5.91 49300742 355.69
5.95 58836310 416.02
6.14 78387147 456.36
6.57 72330158 432.03

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101019\
 Data File : PQ044207.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2019 09:44
 Operator : HP\AJ
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_Q
 LabSampleId :
 AR1248310

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 02:35:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 10 02:55:10 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Manual Integrations
 APPROVED

Ankita
 10/11/2019 1:45:05 PM

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.321	4.357	89509863	101.2E6	23.024	22.681
2) SA Decachlor...	11.610	9.806	190.4E6	215.2E6	35.548	35.976
Target Compounds						
21) L5 AR-1248-1	6.643	5.639	34637964	38456031	472.056	424.684
22) L5 AR-1248-2	6.939	5.905	46049669	49300742	464.866	355.692
23) L5 AR-1248-3	7.160	5.950	49442295	58836310	457.805	416.024
24) L5 AR-1248-4	7.589	6.140	56982836	78387147	452.597	456.362m
25) L5 AR-1248-5	7.630	6.567	57536778	72330158	475.203	432.029m

AJ
 10/15/19

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.