

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101022\
 Data File : PQ059560.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2022 18:15
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR16603017

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 18:32:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100422CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 06 10:42:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.483	2.829	261.8E6	184.0E6	19.010	20.312
2) SA Decachlor...	8.653	7.606	214.5E6	320.0E6	43.193	41.905
Target Compounds						
3) L1 AR-1016-1	4.583	3.837	162.5E6	112.5E6	387.569	398.835
4) L1 AR-1016-2	4.602	3.853	243.7E6	170.7E6	393.030	401.820
5) L1 AR-1016-3	4.660	4.014	157.3E6	89642663	398.624	403.042
6) L1 AR-1016-4	4.752	4.061	129.5E6	69786768	393.154	408.197
7) L1 AR-1016-5	5.038	4.258	121.6E6	91700964	387.646	407.023
31) L7 AR-1260-1	6.136	5.251	235.1E6	182.9E6	422.531	405.852
32) L7 AR-1260-2	6.394	5.439	269.2E6	221.4E6	421.622	402.581
33) L7 AR-1260-3	6.746	5.583	191.0E6	210.5E6	427.134	405.785
34) L7 AR-1260-4	6.964	6.045	203.9E6	177.5E6	422.964	401.587
35) L7 AR-1260-5	7.273	6.288	363.9E6	414.9E6	412.797	406.813

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101022\
Data File : PQ059560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2022 18:15
Operator : YP\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR16603017

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 18:32:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100422CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 06 10:42:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

