

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101022\
 Data File : PQ059591.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Oct 2022 01:51
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR16603021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 04:54:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100422CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 06 10:42:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.484	2.829	253.4E6	185.3E6	18.406	20.453
2) SA Decachlor...	8.655	7.605	206.4E6	322.3E6	41.571	42.213
Target Compounds						
3) L1 AR-1016-1	4.583	3.836	163.4E6	116.3E6	389.586	412.560
4) L1 AR-1016-2	4.603	3.852	238.8E6	175.0E6	385.241	412.024
5) L1 AR-1016-3	4.661	4.014	150.7E6	91845457	381.851	412.946
6) L1 AR-1016-4	4.753	4.061	124.7E6	70226559	378.628	410.770
7) L1 AR-1016-5	5.039	4.257	120.3E6	93281492	383.441	414.038
31) L7 AR-1260-1	6.137	5.251	223.8E6	185.9E6	402.334	412.498
32) L7 AR-1260-2	6.395	5.439	257.9E6	226.9E6	404.040	412.536
33) L7 AR-1260-3	6.747	5.582	179.0E6	213.2E6	400.150	411.047
34) L7 AR-1260-4	6.965	6.044	192.9E6	180.8E6	400.117	409.008
35) L7 AR-1260-5	7.273	6.287	351.3E6	429.1E6	398.413	420.762

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101022\
Data File : PQ059591.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Oct 2022 01:51
Operator : YP\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR16603021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 04:54:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100422CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 06 10:42:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

