

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ071006.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 20:36
 Operator : YP\AJ
 Sample : DDT ANALOG
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 DDT ANALOG

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 01:57:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:55:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
30) L6 AR-1254-5	6.565	5.692	2293.3E6	3164.2E6	9889.519	7527.648



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071006.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 20:36
Operator : YP\AJ
Sample : DDT ANALOG
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
DDT ANALOG

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:57:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 3.400 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
DDT ANALOG




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 2.778 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.496 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 7.533 min
Response: 0
Conc: N.D.



#30 AR-1254-5

R.T.: 6.565 min
Delta R.T.: 0.022 min
Response: 2293336906
Conc: 9889.52 ng/m

Instrument :
ECD_Q
ClientSampleId :
DDT ANALOG



#30 AR-1254-5

R.T.: 5.692 min
Delta R.T.: 0.003 min
Response: 3164243339
Conc: 7527.65 ng/ml