

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101118\
 Data File : PQ032908.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Oct 2018 03:41
 Operator : SM\SJ
 Sample : J5375-03
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 0691-KM-PA-AMD-COMP

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 04:17:54 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 19 02:07:33 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.577	3.857	41240481	32893169	17.786	16.424
2) SA Decachlor...	10.420	8.917	16981432	14356432	8.146	7.414

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101118\
Data File : PQ032908.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2018 03:41
Operator : SM\SJ
Sample : J5375-03
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
0691-KM-PA-AMD-COMP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 12 04:17:54 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ091818.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 19 02:07:33 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.577 min
Delta R.T.: -0.017 min
Response: 41240481
Conc: 17.79 ng/ml

Instrument :
ECD_Q
ClientSampleId :
0691-KM-PA-AMD-COMP




#1 Tetrachloro-m-xylene

R.T.: 3.857 min
Delta R.T.: -0.017 min
Response: 32893169
Conc: 16.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.420 min
Delta R.T.: -0.032 min
Response: 16981432
Conc: 8.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.917 min
Delta R.T.: -0.035 min
Response: 14356432
Conc: 7.41 ng/ml