

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101123\
 Data File : PQ063658.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2023 20:01
 Operator : YP\AJ
 Sample : 04795-07MS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 A0B92MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/11/2023
 Supervised By :Ankita Jodhani 10/11/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 20:59:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092523CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 26 06:41:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.460	2.779	87168206	72646812	16.614	17.976
2)	SA Decachlor...	8.695	7.580	53559086	62768210	14.496	16.976
Target Compounds							
3)	L1 AR-1016-1	4.575	3.790	44652210	46610229	270.345m	350.624 #
4)	L1 AR-1016-2	4.594	3.806	80371906	68178902	319.002m	355.411
5)	L1 AR-1016-3	4.653	3.966	48774269	37224009	311.288m	354.610
6)	L1 AR-1016-4	4.745	4.016	48234982	31912390	376.323m	360.701
7)	L1 AR-1016-5	5.036	4.212	42039506	36897053	333.460	343.576
31)	L7 AR-1260-1	6.146	5.212	145.0E6	134.1E6	586.231	626.898
32)	L7 AR-1260-2	6.407	5.403	260.7E6	227.0E6	873.748	894.795
33)	L7 AR-1260-3	6.765	5.544	146.9E6	116.6E6	671.759m	481.518 #
34)	L7 AR-1260-4	6.982	6.010	134.3E6	113.4E6	545.937m	559.613
35)	L7 AR-1260-5	7.296	6.256	256.4E6	266.9E6	538.282	575.395m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101123\
Data File : PQ063658.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2023 20:01
Operator : YP\AJ
Sample : 04795-07MS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
A0B92MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/11/2023
Supervised By :Ankita Jodhani 10/11/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 20:59:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ092523CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 26 06:41:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

