Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101318\
Data File : PQ033093.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2018 22:58
Operator : SM\SJ
Sample : MDL-AR1232-W-3
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
MDL-AR1232-W-3

Manual Integrations APPROVED
Sohil
10/16/2018 9:16:37 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 05:21:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100918CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 10 08:14:42 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.574	3.853	55463895	41787767	27.036	26.867
2)	SA Decachlor...	10.409	8.908	80778730	68511649	42.733	42.618
	Target Compounds						
11)	L3 AR-1232-1	4.948	4.232	1338389	987939	30.148	28.549
12)	L3 AR-1232-2	5.480	4.964	800906	969328	33.910	25.778
13)	L3 AR-1232-3	5.771	5.143	1270094	561709	26.566	28.679
14)	L3 AR-1232-4	5.936	5.224	847674	416767	35.588m	24.283 #
15)	L3 AR-1232-5	6.024	5.399	610483	568207	34.231	29.139

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101318\
Data File : PQ033093.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2018 22:58
Operator : SM\SJ
Sample : MDL-AR1232-W-3
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
MDL-AR1232-W-3

Manual Integrations APPROVED
Sohil
10/16/2018 9:16:37 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 05:21:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100918CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 10 08:14:42 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm