

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101322\
 Data File : PQ059613.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Oct 2022 09:40
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR16603027

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 13 15:26:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100422CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 06 10:42:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.483	2.829	241.3E6	170.3E6	17.528	18.798
2) SA Decachlor...	8.652	7.605	194.2E6	297.8E6	39.108	39.002
Target Compounds						
3) L1 AR-1016-1	4.582	3.835	151.9E6	105.3E6	362.136	373.271
4) L1 AR-1016-2	4.602	3.851	222.8E6	156.0E6	359.350	367.363
5) L1 AR-1016-3	4.660	4.013	143.9E6	83055267	364.690	373.425
6) L1 AR-1016-4	4.752	4.060	119.9E6	65874862	364.070	385.316
7) L1 AR-1016-5	5.037	4.256	122.1E6	86618529	389.379	384.464
31) L7 AR-1260-1	6.136	5.250	221.1E6	173.3E6	397.431	384.586
32) L7 AR-1260-2	6.392	5.438	252.6E6	209.3E6	395.687	380.585
33) L7 AR-1260-3	6.746	5.581	177.1E6	198.7E6	395.971	383.099
34) L7 AR-1260-4	6.963	6.043	190.3E6	170.3E6	394.642	385.273
35) L7 AR-1260-5	7.272	6.287	331.3E6	391.3E6	375.820	383.717

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101322\
Data File : PQ059613.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Oct 2022 09:40
Operator : YP\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR16603027

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 13 15:26:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100422CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 06 10:42:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

