

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101419\
 Data File : PQ044366.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2019 00:28
 Operator : HP\AJ
 Sample : AIBLK04
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AIBLK04

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 02:53:14 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 10 02:55:10 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.322	4.355	85480557	104.5E6	21.987	23.402
2)	SA Decachlor...	11.602	9.794	247.4E6	296.8E6	46.170	49.625
Target Compounds							
30)	L6 AR-1254-5	8.890	0.000	4967592	0	20.755	N.D. #
33)	L7 AR-1260-3	8.935	0.000	2511350	0	15.663	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101419\
Data File : PQ044366.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2019 00:28
Operator : HP\AJ
Sample : AIBLK04
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AIBLK04

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 02:53:14 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 10 02:55:10 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.322 min
Delta R.T.: 0.005 min
Response: 85480557
Conc: 21.99 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AIBLK04




#1 Tetrachloro-m-xylene

R.T.: 4.355 min
Delta R.T.: 0.000 min
Response: 104464340
Conc: 23.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.602 min
Delta R.T.: 0.003 min
Response: 247358548
Conc: 46.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.794 min
Delta R.T.: -0.014 min
Response: 296824741
Conc: 49.62 ng/ml



#30 AR-1254-5


R.T.: 8.890 min
Delta R.T.: -0.011 min
Response: 4967592
Conc: 20.75 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AIBLK04




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.785 min
Response: 0
Conc: N.D.



#33 AR-1260-3

R.T.: 8.935 min
Delta R.T.: -0.045 min
Response: 2511350
Conc: 15.66 ng/ml



#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 7.606 min
Response: 0
Conc: N.D.