

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101419\
Data File : PQ044330.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2019 13:40
Operator : HP\AJ
Sample : K5343-01DL 500X
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
C0AY3DL

Manual Integrations
APPROVED

Ankita
10/15/2019 8:29:28 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 14:13:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 10 02:55:10 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
7.56 631908846 4696.20
7.79 1101148278 5154.57
8.18 1453862514 5908.86
8.48 1218220109 6074.12
8.91 1740823687 7273.29

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.52 1463276981 5016.95
6.68 1477082212 5657.18
7.11 2864365716 6556.01
7.35 1895513636 7056.34
7.78 3034580481 7589.05

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101419\
Data File : PQ044330.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2019 13:40
Operator : HP\AJ
Sample : K5343-01DL 500X
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
C0AY3DL

Manual Integrations
APPROVED

Ankita
10/15/2019 8:29:28 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 14:13:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 10 02:55:10 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
7.56 631908846 4696.20
7.79 1101148278 5154.57
8.18 1453862514 5908.86
8.48 1218220109 6074.12
8.91 1527968742 6383.97

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.52 1463276981 5016.95
6.68 1477082212 5657.18
7.11 2864365716 6556.01
7.35 1895513636 7056.34
7.78 3034580481 7589.05

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101419\
 Data File : PQ044330.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Oct 2019 13:40
 Operator : HP\AJ
 Sample : K5343-01DL 500X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 Client Sampled :
 C0AY3DL

Manual Integrations
 APPROVED

Ankita
 10/15/2019 8:29:28 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 14:13:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 10 02:55:10 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

26) L6 AR-1254-1	7.564	6.520	631.9E6	1463.3E6	4696.196	5016.953
27) L6 AR-1254-2	7.794	6.678	1101.1E6	1477.1E6	5154.575	5657.183
28) L6 AR-1254-3	8.180	7.108	1453.9E6	2864.4E6	5908.859	6556.008
29) L6 AR-1254-4	8.476	7.347	1218.2E6	1895.5E6	6074.124	7056.340
30) L6 AR-1254-5	8.911	7.782	1528.0E6	3034.6E6	6383.967m	7589.048

A. J
 10115119

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.