

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101420\
 Data File : PQ050399.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Oct 2020 20:58
 Operator : DD\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 02:51:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 14 02:48:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.205	4.269	931.8E6	294.1E6	77.806	78.627
2)	SA Decachlor...	11.395	9.620	614.8E6	495.6E6	71.855	75.222
Target Compounds							
21)	L5 AR-1248-1	6.538	5.530	201.7E6	71311833	743.549	718.602
22)	L5 AR-1248-2	6.831	5.793	276.7E6	100.5E6	750.602	758.957
23)	L5 AR-1248-3	7.050	5.839	305.4E6	104.7E6	749.612	761.775
24)	L5 AR-1248-4	7.482	6.024	356.4E6	128.5E6	742.712	761.811
25)	L5 AR-1248-5	7.521	6.446	344.1E6	145.5E6	740.668	759.358

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101420\
Data File : PQ050399.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Oct 2020 20:58
Operator : DD\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 02:51:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 14 02:48:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

