

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101521\
 Data File : PQ055002.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2021 19:53
 Operator : AJ\MA
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR16603153

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 05:13:54 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100421CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 05 07:08:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.193	4.198	528.3E6	232.8E6	22.810	21.251
2)	SA Decachlor...	11.352	9.523	798.3E6	590.0E6	44.637	41.693
Target Compounds							
3)	L1 AR-1016-1	6.517	5.459	310.4E6	147.9E6	440.294	405.311
4)	L1 AR-1016-2	6.542	5.478	495.2E6	267.6E6	432.480	405.544
5)	L1 AR-1016-3	6.608	5.671	295.7E6	121.2E6	432.247	407.223
6)	L1 AR-1016-4	6.718	5.720	235.7E6	100.1E6	435.751	395.605
7)	L1 AR-1016-5	7.027	5.951	224.1E6	116.5E6	424.800	365.386
31)	L7 AR-1260-1	8.201	7.044	375.6E6	258.5E6	429.415	397.598
32)	L7 AR-1260-2	8.464	7.241	467.1E6	358.2E6	417.192	398.679
33)	L7 AR-1260-3	8.759	7.395	518.5E6	316.8E6	423.623	388.022
34)	L7 AR-1260-4	9.072	7.877	381.2E6	271.5E6	416.780	396.167
35)	L7 AR-1260-5	9.415	8.124	767.6E6	781.0E6	424.768	406.010

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101521\
Data File : PQ055002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2021 19:53
Operator : AJ\MA
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR16603153

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 05:13:54 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100421CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 05 07:08:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

