

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101524\
 Data File : PQ069133.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2024 17:49
 Operator : YP\AJ
 Sample : P4346-07
 Misc :
 ALS Vial : 24 Sample Multiplier: 1



Instrument :
 ECD_Q
 ClientSampleId :
 EOAW9

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 10/17/2024
 Supervised By :Ankita Jodhani 10/17/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 03:07:29 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100724CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 08 08:10:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) AR-1242-1 (L4)
 R.T. Response Conc
 4.42 697637 2.24
 4.44 7725463 16.50
 4.50 8631140 30.33
 4.61 6738771 28.33
 5.30 51777446 208.28

(16) AR-1242-1 #2 (L4)
 R.T. Response Conc
 3.73 6944656 21.06
 3.75 6621416 13.86
 3.91 6462353 25.08
 3.99 14624178 63.03
 4.48 38989065 121.95

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101524\
 Data File : PQ069133.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2024 17:49
 Operator : YP\AJ
 Sample : P4346-07
 Misc :
 ALS Vial : 24 Sample Multiplier: 1



Instrument :
 ECD_Q
 ClientSampleId :
 EOAW9

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 10/17/2024
 Supervised By :Ankita Jodhani 10/17/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 03:07:29 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100724CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 08 08:10:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) AR-1242-1 #2 (L4)
 R.T. Response Conc
 4.42 3171509 10.17
 4.44 9582786 20.47
 4.50 6196618 21.77
 4.61 6738771 28.33
 5.29 11493306 46.23

(16) AR-1242-1 #2 (L4)
 R.T. Response Conc
 3.73 6944656 21.06
 3.75 6621416 13.86
 3.91 4384849 17.02
 3.99 10154407 43.76
 4.48 18245114 57.07

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101524\
 Data File : PQ069133.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2024 17:49
 Operator : YP\AJ
 Sample : P4346-07
 Misc :
 ALS Vial : 24 Sample Multiplier: 1



Instrument :
 ECD_Q
 ClientSampleId :
 EOAW9

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 10/17/2024
 Supervised By :Ankita Jodhani 10/17/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 03:07:29 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100724CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 08 08:10:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 (L6)
 R.T. Response Conc
 5.23 33523723 77.66
 5.44 53387770 79.41
 5.80 70517306 96.84
 6.08 33324202 63.05
 6.49 45201815 77.16

(26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 4.48 38989065 59.21
 4.63 44474404 77.17
 5.01 36476847 39.52
 5.23 27662737 45.96
 5.64 22130947 26.20

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101524\
 Data File : PQ069133.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2024 17:49
 Operator : YP\AJ
 Sample : P4346-07
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :

ECD_Q

ClientSampleId :


EOAW9

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 10/17/2024
 Supervised By :Ankita Jodhani 10/17/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 03:07:29 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100724CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 08 08:10:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)

R.T.	Response	Conc
5.23	11442590	26.51
5.44	10068260	14.98
5.80	24837590	34.11
6.08	20145373	38.12
6.49	21177930	36.15

(26) AR-1254-1 #2 (L6)

R.T.	Response	Conc
4.48	18840393	28.61
4.63	17097582	29.67
5.01	14863300	16.11
5.23	14993193	24.91
5.64	22130947	26.20