

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102124\
 Data File : PQ069218.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2024 20:51
 Operator : YP\AJ
 Sample : P4368-07
 Misc : AR1262 LOD 25 PPB
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 01:52:07 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ102124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 21 15:35:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.329	2.740	185.1E6	190.3E6	20.517	21.084
2)	SA Decachlor...	8.453	7.478	125.7E6	177.3E6	19.466	21.430
Target Compounds							
3)	L1 AR-1016-1	4.417	3.721	254032	618464	0.990	2.195 #
4)	L1 AR-1016-2	4.434	3.740	1155710	1680163	2.670	4.016 #
5)	L1 AR-1016-3	4.493	3.897	926674	3143842	3.474	13.823 #
6)	L1 AR-1016-4	4.582	3.950	552141	1713877	2.452	9.482 #
7)	L1 AR-1016-5	4.869	4.161	920469	1348468	4.337	5.589 #
8)	L2 AR-1221-1	3.526	2.937	851498	970126	8.412	9.593
9)	L2 AR-1221-2	3.605	3.017	1047348	749474	15.995	11.176 #
10)	L2 AR-1221-3	3.670	3.084	637277	983931	3.084	4.498 #
11)	L3 AR-1232-1	3.670	3.076	637277	1415508	4.070	8.505 #
12)	L3 AR-1232-2	4.158	3.740	687655	1680163	6.926	8.925 #
13)	L3 AR-1232-3	4.434	3.897	1155710	3143842	5.755	31.801 #
14)	L3 AR-1232-4	4.582	3.987	552141	1947489	5.393	23.684 #
15)	L3 AR-1232-5	4.677	4.161	799507	1348468	12.615	13.614
16)	L4 AR-1242-1	4.417	3.721	254032	618464	1.063	2.530 #
17)	L4 AR-1242-2	4.434	3.740	1155710	1680163	3.157	4.699 #
18)	L4 AR-1242-3	4.493	3.897	926674	3143842	4.179	15.088 #
19)	L4 AR-1242-4	4.582	3.987	552141	1947489	2.894	11.136 #
20)	L4 AR-1242-5	5.289	4.478	269314	8990234	1.691	34.868 #
21)	L5 AR-1248-1	4.414	3.721	314000	618464	1.621	3.324 #
22)	L5 AR-1248-2	4.677	3.950	799507	1713877	3.218	6.472 #
23)	L5 AR-1248-3	4.869	3.987	920469	1947489	3.050	7.830 #
24)	L5 AR-1248-4	5.259	4.161	1921993	1348468	5.473	4.002 #
25)	L5 AR-1248-5	5.289	4.523	269314	1994802	0.799	5.778 #
26)	L6 AR-1254-1	5.232	4.478	5331245	8990234	15.259	17.063
27)	L6 AR-1254-2	5.446	4.625	1766727	5230353	3.189	10.543 #
28)	L6 AR-1254-3	5.794	5.006	1431375	6037258	2.315	7.883 #
29)	L6 AR-1254-4	6.080	5.239	590624	3862967	1.339	8.281 #
30)	L6 AR-1254-5	6.494	5.632	4916490	14172918	9.999	25.902 #
31)	L7 AR-1260-1	5.963	5.138	7045183	12524505	17.310	27.826 #
32)	L7 AR-1260-2	6.220	5.325	10573972	11758690	21.357	23.286
33)	L7 AR-1260-3	6.573	5.464	6770939	9020738	18.464	17.804
34)	L7 AR-1260-4	6.792	5.924	6697122	15883681	15.888	35.954 #
35)	L7 AR-1260-5	7.101	6.170	25056038	33305754	29.403	32.645
36)	L8 AR-1262-1	6.788	5.678	6221944	21131151	11.257	31.452 #
37)	L8 AR-1262-2	7.101	5.924	25056038	15883681	23.757	25.538
38)	L8 AR-1262-3	7.374	6.444	9845128	32662945	14.379	60.178 #
39)	L8 AR-1262-4	7.450	6.507	3651220	34994232	7.000	37.585 #
40)	L8 AR-1262-5	7.950	6.998	5228266	7019957	15.622	14.678
41)	L9 AR-1268-1	7.374	6.444	9845128	32662945	8.476	23.005 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102124\
Data File : PQ069218.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2024 20:51
Operator : YP\AJ
Sample : P4368-07
Misc : AR1262 LOD 25 PPB
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 01:52:07 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ102124.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 21 15:35:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	7.450	6.507	3651220	34994232	3.417	27.294 #
43) L9	AR-1268-3	7.629	6.707	859452	2643363	0.941	2.426 #
44) L9	AR-1268-4	7.950	6.998	5228266	7019957	13.279	16.061
45) L9	AR-1268-5	8.234	7.262	2272438	7304084	0.871	2.162 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102124\
Data File : PQ069218.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2024 20:51
Operator : YP\AJ
Sample : P4368-07
Misc : AR1262 LOD 25 PPB
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 01:52:07 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ102124.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 21 15:35:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.329 min
Delta R.T.: 0.000 min
Response: 185122600
Conc: 20.52 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#1 Tetrachloro-m-xylene

R.T.: 2.740 min
Delta R.T.: 0.001 min
Response: 190343433
Conc: 21.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.453 min
Delta R.T.: 0.000 min
Response: 125731390
Conc: 19.47 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.478 min
Delta R.T.: 0.000 min
Response: 177290884
Conc: 21.43 ng/ml



#3 AR-1016-1


R.T.: 4.417 min
Delta R.T.: 0.001 min
Response: 254032
Conc: 0.99 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#3 AR-1016-1

R.T.: 3.721 min
Delta R.T.: -0.009 min
Response: 618464
Conc: 2.19 ng/ml




#4 AR-1016-2

R.T.: 4.434 min
Delta R.T.: -0.001 min
Response: 1155710
Conc: 2.67 ng/ml



#4 AR-1016-2


R.T.: 3.740 min
Delta R.T.: -0.005 min
Response: 1680163
Conc: 4.02 ng/ml



#5 AR-1016-3


R.T.: 4.493 min
Delta R.T.: 0.000 min
Response: 926674
Conc: 3.47 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#5 AR-1016-3

R.T.: 3.897 min
Delta R.T.: -0.008 min
Response: 3143842
Conc: 13.82 ng/ml




#6 AR-1016-4

R.T.: 4.582 min
Delta R.T.: -0.001 min
Response: 552141
Conc: 2.45 ng/ml



#6 AR-1016-4


R.T.: 3.950 min
Delta R.T.: -0.003 min
Response: 1713877
Conc: 9.48 ng/ml



#7 AR-1016-5


R.T.: 4.869 min
Delta R.T.: 0.002 min
Response: 920469
Conc: 4.34 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#7 AR-1016-5

R.T.: 4.161 min
Delta R.T.: 0.014 min
Response: 1348468
Conc: 5.59 ng/ml




#8 AR-1221-1

R.T.: 3.526 min
Delta R.T.: 0.000 min
Response: 851498
Conc: 8.41 ng/ml



#8 AR-1221-1


R.T.: 2.937 min
Delta R.T.: 0.000 min
Response: 970126
Conc: 9.59 ng/ml



#9 AR-1221-2


R.T.: 3.605 min
Delta R.T.: 0.001 min
Response: 1047348
Conc: 16.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#9 AR-1221-2

R.T.: 3.017 min
Delta R.T.: 0.005 min
Response: 749474
Conc: 11.18 ng/ml




#10 AR-1221-3

R.T.: 3.670 min
Delta R.T.: -0.003 min
Response: 637277
Conc: 3.08 ng/ml



#10 AR-1221-3


R.T.: 3.084 min
Delta R.T.: 0.004 min
Response: 983931
Conc: 4.50 ng/ml



#11 AR-1232-1


R.T.: 3.670 min
Delta R.T.: -0.002 min
Response: 637277
Conc: 4.07 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#11 AR-1232-1

R.T.: 3.076 min
Delta R.T.: -0.004 min
Response: 1415508
Conc: 8.50 ng/ml




#12 AR-1232-2

R.T.: 4.158 min
Delta R.T.: -0.001 min
Response: 687655
Conc: 6.93 ng/ml



#12 AR-1232-2


R.T.: 3.740 min
Delta R.T.: -0.005 min
Response: 1680163
Conc: 8.93 ng/ml



#13 AR-1232-3


R.T.: 4.434 min
Delta R.T.: -0.001 min
Response: 1155710
Conc: 5.76 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#13 AR-1232-3

R.T.: 3.897 min
Delta R.T.: -0.008 min
Response: 3143842
Conc: 31.80 ng/ml




#14 AR-1232-4

R.T.: 4.582 min
Delta R.T.: -0.001 min
Response: 552141
Conc: 5.39 ng/ml



#14 AR-1232-4


R.T.: 3.987 min
Delta R.T.: -0.003 min
Response: 1947489
Conc: 23.68 ng/ml



#15 AR-1232-5


R.T.: 4.677 min
Delta R.T.: 0.000 min
Response: 799507
Conc: 12.61 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#15 AR-1232-5

R.T.: 4.161 min
Delta R.T.: 0.014 min
Response: 1348468
Conc: 13.61 ng/ml




#16 AR-1242-1

R.T.: 4.417 min
Delta R.T.: 0.000 min
Response: 254032
Conc: 1.06 ng/ml



#16 AR-1242-1


R.T.: 3.721 min
Delta R.T.: -0.009 min
Response: 618464
Conc: 2.53 ng/ml



#17 AR-1242-2


R.T.: 4.434 min
Delta R.T.: 0.000 min
Response: 1155710
Conc: 3.16 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#17 AR-1242-2

R.T.: 3.740 min
Delta R.T.: -0.006 min
Response: 1680163
Conc: 4.70 ng/ml




#18 AR-1242-3

R.T.: 4.493 min
Delta R.T.: 0.001 min
Response: 926674
Conc: 4.18 ng/ml



#18 AR-1242-3


R.T.: 3.897 min
Delta R.T.: -0.008 min
Response: 3143842
Conc: 15.09 ng/ml



#19 AR-1242-4


R.T.: 4.582 min
Delta R.T.: 0.000 min
Response: 552141
Conc: 2.89 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#19 AR-1242-4

R.T.: 3.987 min
Delta R.T.: -0.003 min
Response: 1947489
Conc: 11.14 ng/ml




#20 AR-1242-5

R.T.: 5.289 min
Delta R.T.: -0.003 min
Response: 269314
Conc: 1.69 ng/ml



#20 AR-1242-5


R.T.: 4.478 min
Delta R.T.: -0.005 min
Response: 8990234
Conc: 34.87 ng/ml



#21 AR-1248-1


R.T.: 4.414 min
Delta R.T.: -0.001 min
Response: 314000
Conc: 1.62 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#21 AR-1248-1

R.T.: 3.721 min
Delta R.T.: -0.009 min
Response: 618464
Conc: 3.32 ng/ml




#22 AR-1248-2

R.T.: 4.677 min
Delta R.T.: 0.000 min
Response: 799507
Conc: 3.22 ng/ml



#22 AR-1248-2


R.T.: 3.950 min
Delta R.T.: -0.003 min
Response: 1713877
Conc: 6.47 ng/ml



#23 AR-1248-3


R.T.: 4.869 min
Delta R.T.: 0.004 min
Response: 920469
Conc: 3.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#23 AR-1248-3

R.T.: 3.987 min
Delta R.T.: -0.004 min
Response: 1947489
Conc: 7.83 ng/ml




#24 AR-1248-4

R.T.: 5.259 min
Delta R.T.: 0.002 min
Response: 1921993
Conc: 5.47 ng/ml



#24 AR-1248-4


R.T.: 4.161 min
Delta R.T.: 0.014 min
Response: 1348468
Conc: 4.00 ng/ml



#25 AR-1248-5


R.T.: 5.289 min
Delta R.T.: -0.004 min
Response: 269314
Conc: 0.80 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#25 AR-1248-5

R.T.: 4.523 min
Delta R.T.: 0.004 min
Response: 1994802
Conc: 5.78 ng/ml




#26 AR-1254-1

R.T.: 5.232 min
Delta R.T.: 0.003 min
Response: 5331245
Conc: 15.26 ng/ml



#26 AR-1254-1


R.T.: 4.478 min
Delta R.T.: -0.003 min
Response: 8990234
Conc: 17.06 ng/ml



#27 AR-1254-2


R.T.: 5.446 min
Delta R.T.: 0.000 min
Response: 1766727
Conc: 3.19 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#27 AR-1254-2

R.T.: 4.625 min
Delta R.T.: -0.002 min
Response: 5230353
Conc: 10.54 ng/ml




#28 AR-1254-3

R.T.: 5.794 min
Delta R.T.: -0.004 min
Response: 1431375
Conc: 2.31 ng/ml



#28 AR-1254-3


R.T.: 5.006 min
Delta R.T.: 0.000 min
Response: 6037258
Conc: 7.88 ng/ml



#29 AR-1254-4


R.T.: 6.080 min
Delta R.T.: -0.001 min
Response: 590624
Conc: 1.34 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#29 AR-1254-4

R.T.: 5.239 min
Delta R.T.: 0.007 min
Response: 3862967
Conc: 8.28 ng/ml




#30 AR-1254-5

R.T.: 6.494 min
Delta R.T.: 0.002 min
Response: 4916490
Conc: 10.00 ng/ml



#30 AR-1254-5


R.T.: 5.632 min
Delta R.T.: -0.001 min
Response: 14172918
Conc: 25.90 ng/ml



#31 AR-1260-1


R.T.: 5.963 min
Delta R.T.: 0.000 min
Response: 7045183
Conc: 17.31 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#31 AR-1260-1

R.T.: 5.138 min
Delta R.T.: 0.003 min
Response: 12524505
Conc: 27.83 ng/ml




#32 AR-1260-2

R.T.: 6.220 min
Delta R.T.: -0.001 min
Response: 10573972
Conc: 21.36 ng/ml



#32 AR-1260-2


R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 11758690
Conc: 23.29 ng/ml



#33 AR-1260-3


R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 6770939
Conc: 18.46 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#33 AR-1260-3

R.T.: 5.464 min
Delta R.T.: -0.001 min
Response: 9020738
Conc: 17.80 ng/ml




#34 AR-1260-4

R.T.: 6.792 min
Delta R.T.: 0.002 min
Response: 6697122
Conc: 15.89 ng/ml



#34 AR-1260-4


R.T.: 5.924 min
Delta R.T.: 0.002 min
Response: 15883681
Conc: 35.95 ng/ml



#35 AR-1260-5


R.T.: 7.101 min
Delta R.T.: 0.000 min
Response: 25056038
Conc: 29.40 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#35 AR-1260-5

R.T.: 6.170 min
Delta R.T.: 0.000 min
Response: 33305754
Conc: 32.64 ng/ml




#36 AR-1262-1

R.T.: 6.788 min
Delta R.T.: -0.001 min
Response: 6221944
Conc: 11.26 ng/ml



#36 AR-1262-1


R.T.: 5.678 min
Delta R.T.: 0.002 min
Response: 21131151
Conc: 31.45 ng/ml



#37 AR-1262-2


R.T.: 7.101 min
Delta R.T.: 0.000 min
Response: 25056038
Conc: 23.76 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#37 AR-1262-2

R.T.: 5.924 min
Delta R.T.: 0.003 min
Response: 15883681
Conc: 25.54 ng/ml




#38 AR-1262-3

R.T.: 7.374 min
Delta R.T.: 0.000 min
Response: 9845128
Conc: 14.38 ng/ml



#38 AR-1262-3


R.T.: 6.444 min
Delta R.T.: 0.000 min
Response: 32662945
Conc: 60.18 ng/ml



#39 AR-1262-4


R.T.: 7.450 min
Delta R.T.: 0.002 min
Response: 3651220
Conc: 7.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#39 AR-1262-4

R.T.: 6.507 min
Delta R.T.: 0.002 min
Response: 34994232
Conc: 37.58 ng/ml




#40 AR-1262-5

R.T.: 7.950 min
Delta R.T.: 0.000 min
Response: 5228266
Conc: 15.62 ng/ml



#40 AR-1262-5


R.T.: 6.998 min
Delta R.T.: 0.002 min
Response: 7019957
Conc: 14.68 ng/ml



#41 AR-1268-1


R.T.: 7.374 min
Delta R.T.: 0.001 min
Response: 9845128
Conc: 8.48 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#41 AR-1268-1

R.T.: 6.444 min
Delta R.T.: 0.000 min
Response: 32662945
Conc: 23.01 ng/ml




#42 AR-1268-2

R.T.: 7.450 min
Delta R.T.: 0.000 min
Response: 3651220
Conc: 3.42 ng/ml



#42 AR-1268-2


R.T.: 6.507 min
Delta R.T.: 0.000 min
Response: 34994232
Conc: 27.29 ng/ml



#43 AR-1268-3


R.T.: 7.629 min
Delta R.T.: -0.001 min
Response: 859452
Conc: 0.94 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#43 AR-1268-3

R.T.: 6.707 min
Delta R.T.: 0.002 min
Response: 2643363
Conc: 2.43 ng/ml




#44 AR-1268-4

R.T.: 7.950 min
Delta R.T.: -0.002 min
Response: 5228266
Conc: 13.28 ng/ml



#44 AR-1268-4


R.T.: 6.998 min
Delta R.T.: 0.002 min
Response: 7019957
Conc: 16.06 ng/ml



#45 AR-1268-5

R.T.: 8.234 min
Delta R.T.: 0.003 min
Response: 2272438
Conc: 0.87 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024



#45 AR-1268-5

R.T.: 7.262 min
Delta R.T.: 0.000 min
Response: 7304084
Conc: 2.16 ng/ml