

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102124\
 Data File : PQ069219.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2024 21:06
 Operator : YP\AJ
 Sample : P4368-07
 Misc : AR1268 LOD 25 PPB
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 01:52:40 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ102124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 21 15:35:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.328	2.739	159.0E6	170.7E6	17.625	18.912
2)	SA Decachlor...	8.454	7.477	132.5E6	187.4E6	20.509	22.647
Target Compounds							
3)	L1 AR-1016-1	4.425	3.723	3419202	313950	13.325	1.114 #
4)	L1 AR-1016-2	4.425	3.741	3419202	2191248	7.900	5.238 #
5)	L1 AR-1016-3	4.488	3.911	9568730	1489793	35.876	6.550 #
6)	L1 AR-1016-4	4.565	3.935	4174635	2595917	18.540	14.362
7)	L1 AR-1016-5	4.875	4.144	4626308	3020629	21.799	12.519 #
8)	L2 AR-1221-1	3.519	2.940	2630712	399027	25.989	3.946 #
9)	L2 AR-1221-2	3.603	3.013	7037374	1459650	107.475	21.766 #
10)	L2 AR-1221-3	3.680	3.078	1513247	2279321	7.324	10.420 #
11)	L3 AR-1232-1	3.665	3.078	1058859	2279321	6.762	13.694 #
12)	L3 AR-1232-2	4.160	3.741	1908677	2191248	19.223	11.640 #
13)	L3 AR-1232-3	4.425	3.911	3419202	1489793	17.027	15.070
14)	L3 AR-1232-4	4.565	3.992	4174635	1980027	40.777	24.080 #
15)	L3 AR-1232-5	4.666	4.144	2890945	3020629	45.614	30.496 #
16)	L4 AR-1242-1	4.425	3.723	3419202	313950	14.302	1.284 #
17)	L4 AR-1242-2	4.425	3.741	3419202	2191248	9.339	6.129 #
18)	L4 AR-1242-3	4.488	3.911	9568730	1489793	43.151	7.150 #
19)	L4 AR-1242-4	4.565	3.992	4174635	1980027	21.881	11.322 #
20)	L4 AR-1242-5	5.272	4.498	2226844	3686220	13.986	14.297
21)	L5 AR-1248-1	4.425	3.723	3419202	313950	17.654	1.688 #
22)	L5 AR-1248-2	4.666	3.935	2890945	2595917	11.637	9.802
23)	L5 AR-1248-3	4.875	3.992	4626308	1980027	15.328	7.961 #
24)	L5 AR-1248-4	5.272	4.144	2226844	3020629	6.341	8.964 #
25)	L5 AR-1248-5	5.272	4.520	2226844	6348717	6.609	18.390 #
26)	L6 AR-1254-1	5.219	4.498	13160412	3686220	37.666	6.996 #
27)	L6 AR-1254-2	5.490f	4.619	16537018	28447035	29.851	57.340 #
28)	L6 AR-1254-3	5.777	4.984	33114643	11863070	53.547	15.489 #
29)	L6 AR-1254-4	0.000	5.229	0	5258185	N.D.	11.273 #
30)	L6 AR-1254-5	6.486	5.623	4458243	1865401	9.067	3.409 #
31)	L7 AR-1260-1	5.966	5.105f	6499448	10729263	15.969	23.838 #
32)	L7 AR-1260-2	0.000	5.325	0	770121	N.D.	1.525 #
33)	L7 AR-1260-3	6.572	5.427f	10520887	3960969	28.690	7.817 #
34)	L7 AR-1260-4	6.810	5.921	14987885	13477264	35.556	30.507
35)	L7 AR-1260-5	7.099	6.166	3895892	14415216	4.572	14.129 #
36)	L8 AR-1262-1	6.810	5.679	14987885	10831628	27.116	16.122 #
37)	L8 AR-1262-2	7.099	5.921	3895892	13477264	3.694	21.669 #
38)	L8 AR-1262-3	7.371	6.445	24030184	58912714	35.096	108.541 #
39)	L8 AR-1262-4	7.450	6.508	25653765	41813397	49.182	44.909
40)	L8 AR-1262-5	7.952	6.997	14423797	8139049	43.099	17.018 #
41)	L9 AR-1268-1	7.371	6.445	24030184	58912714	20.687	41.494 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102124\
Data File : PQ069219.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2024 21:06
Operator : YP\AJ
Sample : P4368-07
Misc : AR1268 LOD 25 PPB
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 01:52:40 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ102124.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 21 15:35:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	7.450	6.508	25653765	41813397	24.010	32.613 #
43) L9	AR-1268-3	7.629	6.706	21039040	25636989	23.045	23.533
44) L9	AR-1268-4	7.952	6.997	14423797	8139049	36.634	18.621 #
45) L9	AR-1268-5	8.233	7.263	58582665	78806031	22.457	23.330


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102124\
Data File : PQ069219.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2024 21:06
Operator : YP\AJ
Sample : P4368-07
Misc : AR1268 LOD 25 PPB
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 01:52:40 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ102124.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 21 15:35:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.328 min
Delta R.T.: 0.000 min
Response: 159026431
Conc: 17.63 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#1 Tetrachloro-m-xylene

R.T.: 2.739 min
Delta R.T.: 0.000 min
Response: 170733036
Conc: 18.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.454 min
Delta R.T.: 0.001 min
Response: 132467755
Conc: 20.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.477 min
Delta R.T.: 0.000 min
Response: 187357306
Conc: 22.65 ng/ml



#3 AR-1016-1


R.T.: 4.425 min
Delta R.T.: 0.010 min
Response: 3419202
Conc: 13.33 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#3 AR-1016-1

R.T.: 3.723 min
Delta R.T.: -0.008 min
Response: 313950
Conc: 1.11 ng/ml




#4 AR-1016-2

R.T.: 4.425 min
Delta R.T.: -0.010 min
Response: 3419202
Conc: 7.90 ng/ml



#4 AR-1016-2


R.T.: 3.741 min
Delta R.T.: -0.005 min
Response: 2191248
Conc: 5.24 ng/ml



#5 AR-1016-3


R.T.: 4.488 min
Delta R.T.: -0.004 min
Response: 9568730
Conc: 35.88 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#5 AR-1016-3

R.T.: 3.911 min
Delta R.T.: 0.007 min
Response: 1489793
Conc: 6.55 ng/ml




#6 AR-1016-4

R.T.: 4.565 min
Delta R.T.: -0.018 min
Response: 4174635
Conc: 18.54 ng/ml



#6 AR-1016-4


R.T.: 3.935 min
Delta R.T.: -0.018 min
Response: 2595917
Conc: 14.36 ng/ml



#7 AR-1016-5


R.T.: 4.875 min
Delta R.T.: 0.008 min
Response: 4626308
Conc: 21.80 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#7 AR-1016-5

R.T.: 4.144 min
Delta R.T.: -0.003 min
Response: 3020629
Conc: 12.52 ng/ml




#8 AR-1221-1

R.T.: 3.519 min
Delta R.T.: -0.006 min
Response: 2630712
Conc: 25.99 ng/ml



#8 AR-1221-1


R.T.: 2.940 min
Delta R.T.: 0.003 min
Response: 399027
Conc: 3.95 ng/ml



#9 AR-1221-2


R.T.: 3.603 min
Delta R.T.: 0.000 min
Response: 7037374
Conc: 107.47 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#9 AR-1221-2

R.T.: 3.013 min
Delta R.T.: 0.000 min
Response: 1459650
Conc: 21.77 ng/ml




#10 AR-1221-3

R.T.: 3.680 min
Delta R.T.: 0.008 min
Response: 1513247
Conc: 7.32 ng/ml



#10 AR-1221-3


R.T.: 3.078 min
Delta R.T.: -0.002 min
Response: 2279321
Conc: 10.42 ng/ml



#11 AR-1232-1


R.T.: 3.665 min
Delta R.T.: -0.008 min
Response: 1058859
Conc: 6.76 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#11 AR-1232-1

R.T.: 3.078 min
Delta R.T.: -0.002 min
Response: 2279321
Conc: 13.69 ng/ml




#12 AR-1232-2

R.T.: 4.160 min
Delta R.T.: 0.000 min
Response: 1908677
Conc: 19.22 ng/ml



#12 AR-1232-2


R.T.: 3.741 min
Delta R.T.: -0.005 min
Response: 2191248
Conc: 11.64 ng/ml



#13 AR-1232-3


R.T.: 4.425 min
Delta R.T.: -0.009 min
Response: 3419202
Conc: 17.03 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#13 AR-1232-3

R.T.: 3.911 min
Delta R.T.: 0.007 min
Response: 1489793
Conc: 15.07 ng/ml




#14 AR-1232-4

R.T.: 4.565 min
Delta R.T.: -0.018 min
Response: 4174635
Conc: 40.78 ng/ml



#14 AR-1232-4


R.T.: 3.992 min
Delta R.T.: 0.002 min
Response: 1980027
Conc: 24.08 ng/ml



#15 AR-1232-5


R.T.: 4.666 min
Delta R.T.: -0.010 min
Response: 2890945
Conc: 45.61 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#15 AR-1232-5

R.T.: 4.144 min
Delta R.T.: -0.003 min
Response: 3020629
Conc: 30.50 ng/ml




#16 AR-1242-1

R.T.: 4.425 min
Delta R.T.: 0.009 min
Response: 3419202
Conc: 14.30 ng/ml



#16 AR-1242-1


R.T.: 3.723 min
Delta R.T.: -0.008 min
Response: 313950
Conc: 1.28 ng/ml



#17 AR-1242-2


R.T.: 4.425 min
Delta R.T.: -0.009 min
Response: 3419202
Conc: 9.34 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#17 AR-1242-2

R.T.: 3.741 min
Delta R.T.: -0.005 min
Response: 2191248
Conc: 6.13 ng/ml




#18 AR-1242-3

R.T.: 4.488 min
Delta R.T.: -0.004 min
Response: 9568730
Conc: 43.15 ng/ml



#18 AR-1242-3


R.T.: 3.911 min
Delta R.T.: 0.007 min
Response: 1489793
Conc: 7.15 ng/ml



#19 AR-1242-4


R.T.: 4.565 min
Delta R.T.: -0.017 min
Response: 4174635
Conc: 21.88 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#19 AR-1242-4

R.T.: 3.992 min
Delta R.T.: 0.002 min
Response: 1980027
Conc: 11.32 ng/ml




#20 AR-1242-5

R.T.: 5.272 min
Delta R.T.: -0.020 min
Response: 2226844
Conc: 13.99 ng/ml



#20 AR-1242-5


R.T.: 4.498 min
Delta R.T.: 0.015 min
Response: 3686220
Conc: 14.30 ng/ml



#21 AR-1248-1


R.T.: 4.425 min
Delta R.T.: 0.010 min
Response: 3419202
Conc: 17.65 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#21 AR-1248-1

R.T.: 3.723 min
Delta R.T.: -0.007 min
Response: 313950
Conc: 1.69 ng/ml




#22 AR-1248-2

R.T.: 4.666 min
Delta R.T.: -0.012 min
Response: 2890945
Conc: 11.64 ng/ml



#22 AR-1248-2


R.T.: 3.935 min
Delta R.T.: -0.018 min
Response: 2595917
Conc: 9.80 ng/ml



#23 AR-1248-3


R.T.: 4.875 min
Delta R.T.: 0.010 min
Response: 4626308
Conc: 15.33 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#23 AR-1248-3

R.T.: 3.992 min
Delta R.T.: 0.002 min
Response: 1980027
Conc: 7.96 ng/ml




#24 AR-1248-4

R.T.: 5.272 min
Delta R.T.: 0.015 min
Response: 2226844
Conc: 6.34 ng/ml



#24 AR-1248-4


R.T.: 4.144 min
Delta R.T.: -0.002 min
Response: 3020629
Conc: 8.96 ng/ml



#25 AR-1248-5


R.T.: 5.272 min
Delta R.T.: -0.021 min
Response: 2226844
Conc: 6.61 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#25 AR-1248-5

R.T.: 4.520 min
Delta R.T.: 0.001 min
Response: 6348717
Conc: 18.39 ng/ml




#26 AR-1254-1

R.T.: 5.219 min
Delta R.T.: -0.011 min
Response: 13160412
Conc: 37.67 ng/ml



#26 AR-1254-1


R.T.: 4.498 min
Delta R.T.: 0.017 min
Response: 3686220
Conc: 7.00 ng/ml



#27 AR-1254-2


R.T.: 5.490 min
Delta R.T.: 0.044 min
Response: 16537018
Conc: 29.85 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#27 AR-1254-2

R.T.: 4.619 min
Delta R.T.: -0.009 min
Response: 28447035
Conc: 57.34 ng/ml




#28 AR-1254-3

R.T.: 5.777 min
Delta R.T.: -0.021 min
Response: 33114643
Conc: 53.55 ng/ml



#28 AR-1254-3


R.T.: 4.984 min
Delta R.T.: -0.021 min
Response: 11863070
Conc: 15.49 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.081 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#29 AR-1254-4

R.T.: 5.229 min
Delta R.T.: -0.002 min
Response: 5258185
Conc: 11.27 ng/ml




#30 AR-1254-5

R.T.: 6.486 min
Delta R.T.: -0.006 min
Response: 4458243
Conc: 9.07 ng/ml



#30 AR-1254-5


R.T.: 5.623 min
Delta R.T.: -0.011 min
Response: 1865401
Conc: 3.41 ng/ml



#31 AR-1260-1


R.T.: 5.966 min
Delta R.T.: 0.004 min
Response: 6499448
Conc: 15.97 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#31 AR-1260-1

R.T.: 5.105 min
Delta R.T.: -0.029 min
Response: 10729263
Conc: 23.84 ng/ml




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 6.221 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 770121
Conc: 1.53 ng/ml



#33 AR-1260-3


R.T.: 6.572 min
Delta R.T.: -0.002 min
Response: 10520887
Conc: 28.69 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#33 AR-1260-3

R.T.: 5.427 min
Delta R.T.: -0.038 min
Response: 3960969
Conc: 7.82 ng/ml




#34 AR-1260-4

R.T.: 6.810 min
Delta R.T.: 0.020 min
Response: 14987885
Conc: 35.56 ng/ml



#34 AR-1260-4


R.T.: 5.921 min
Delta R.T.: -0.002 min
Response: 13477264
Conc: 30.51 ng/ml



#35 AR-1260-5


R.T.: 7.099 min
Delta R.T.: -0.001 min
Response: 3895892
Conc: 4.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#35 AR-1260-5

R.T.: 6.166 min
Delta R.T.: -0.003 min
Response: 14415216
Conc: 14.13 ng/ml




#36 AR-1262-1

R.T.: 6.810 min
Delta R.T.: 0.021 min
Response: 14987885
Conc: 27.12 ng/ml



#36 AR-1262-1


R.T.: 5.679 min
Delta R.T.: 0.002 min
Response: 10831628
Conc: 16.12 ng/ml



#37 AR-1262-2


R.T.: 7.099 min
Delta R.T.: -0.002 min
Response: 3895892
Conc: 3.69 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#37 AR-1262-2

R.T.: 5.921 min
Delta R.T.: -0.001 min
Response: 13477264
Conc: 21.67 ng/ml




#38 AR-1262-3

R.T.: 7.371 min
Delta R.T.: -0.004 min
Response: 24030184
Conc: 35.10 ng/ml



#38 AR-1262-3


R.T.: 6.445 min
Delta R.T.: 0.002 min
Response: 58912714
Conc: 108.54 ng/ml



#39 AR-1262-4


R.T.: 7.450 min
Delta R.T.: 0.003 min
Response: 25653765
Conc: 49.18 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#39 AR-1262-4

R.T.: 6.508 min
Delta R.T.: 0.003 min
Response: 41813397
Conc: 44.91 ng/ml




#40 AR-1262-5

R.T.: 7.952 min
Delta R.T.: 0.002 min
Response: 14423797
Conc: 43.10 ng/ml



#40 AR-1262-5


R.T.: 6.997 min
Delta R.T.: 0.001 min
Response: 8139049
Conc: 17.02 ng/ml



#41 AR-1268-1


R.T.: 7.371 min
Delta R.T.: -0.002 min
Response: 24030184
Conc: 20.69 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#41 AR-1268-1

R.T.: 6.445 min
Delta R.T.: 0.000 min
Response: 58912714
Conc: 41.49 ng/ml




#42 AR-1268-2

R.T.: 7.450 min
Delta R.T.: 0.000 min
Response: 25653765
Conc: 24.01 ng/ml



#42 AR-1268-2


R.T.: 6.508 min
Delta R.T.: 0.001 min
Response: 41813397
Conc: 32.61 ng/ml



#43 AR-1268-3


R.T.: 7.629 min
Delta R.T.: 0.000 min
Response: 21039040
Conc: 23.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024




#43 AR-1268-3

R.T.: 6.706 min
Delta R.T.: 0.000 min
Response: 25636989
Conc: 23.53 ng/ml




#44 AR-1268-4

R.T.: 7.952 min
Delta R.T.: 0.000 min
Response: 14423797
Conc: 36.63 ng/ml



#44 AR-1268-4


R.T.: 6.997 min
Delta R.T.: 0.000 min
Response: 8139049
Conc: 18.62 ng/ml



#45 AR-1268-5

R.T.: 8.233 min
Delta R.T.: 0.002 min
Response: 58582665
Conc: 22.46 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024



#45 AR-1268-5

R.T.: 7.263 min
Delta R.T.: 0.001 min
Response: 78806031
Conc: 23.33 ng/ml