

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102220\
 Data File : PQ050623.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Oct 2020 00:24
 Operator : DD\AJ
 Sample : L4214-07
 Misc : AR1232 LOD 40 PPB
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2020

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 10:33:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ102120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 22 02:49:52 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.192	4.264	366.7E6	105.2E6	16.862	16.795
2) SA Decachlor...	11.365	9.609	241.1E6	119.2E6	19.279	17.941
Target Compounds						
11) L3 AR-1232-1	5.637	4.709	22134985	4620966	56.697	39.503 #
12) L3 AR-1232-2	6.229	5.543	8566234	5680151	44.340	46.042
13) L3 AR-1232-3	6.548	5.737	18867892	3104225	47.826	48.421
14) L3 AR-1232-4	6.724	5.832	9524899	2593593	48.689	47.965
15) L3 AR-1232-5	6.817	6.017	6056757	2592506	42.828	43.176

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102220\
Data File : PQ050623.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Oct 2020 00:24
Operator : DD\AJ
Sample : L4214-07
Misc : AR1232 LOD 40 PPB
ALS Vial : 62 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2020

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 10:33:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ102120.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 22 02:49:52 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

